



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5362 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: TOXL
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 741 psi Regulator: 750 psi
Display and Regulator \pm 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5321
Lot #: 202311D
Exp. Date: 11/28/25
☒ Results \pm 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP6035

Lot #: 202402C

Exp. Date: 02/14/24

☒ Results \pm 5% AC of known AC

3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 28423080A3

Cylinder #: 41

Exp. Date: 11/5/25

4. ☒ Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS8

Exp. Date: None

☒ Display reads "Interferent Detect"

5. ☒ RFI Check (CMS Mode)

☒ Display reads "RFI Detect"

6. ☒ Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.081 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.081

Test 3 0.080 Test 6 0.081 Test 9 0.080

Average 0.080

☒ Results \pm 0.005 AC of known AC

Post calibration adjust.
0.400 AC
Sim ser #: MP5318
Lot #: 202312D
Exp. date: 12/17/25
☒ Results \pm 5% AC of
Known AC

D. Remarks/Maintenance: calibration adjustment was performed due to
high dry calibration check values (0.083, 0.082, 0.083, 0.083, 0.082 and
0.083 AC). The values listed in '6' are post-calibration adjustment.
aen CG 06 June 2025

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

06 June 2025 aen
05 June 2025
Date

[Signature]
Reviewer

10 June 2025
Date

CMI, Inc. Intoxilyzer	Alcohol Analyzer
North Dakota Model 8000	SN 80-005362
Location = TOXL	8164.14.00 09/16
04/04/2025	12:51

```
abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*()_+?
```

```
abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*() +?
```

Data Entry Mode:	Enabled
Start Test Sequence:	DABACABA
Display Prelim Rslt?	Yes
Display Third Digit?	Yes
Inhib Printer(Y/N)?	No
Display Volume?	No
Disable On Memfull?	Yes
# of Print Copies?	1
Select Std (D/W/I)?	Dry
Standard Value?	0.080
Standard Lot #?	05620080A1
Standard Cyl #?	37
Standard Expiration?	04/05/2022
Oper No?	133237

```
Flow Cal. Date:      05/22/2020
Slope                659
Intercept            -620990
```

IR Calibration Date: 05/26/2020
3um 9um

0th Coef(*100):	-20883	-25370
1st Coef(*100):	265508	144960
2nd Coef(*100):	936	714
H2O adj (mg/l*10k):	391	379

Alan Mil

Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
04/04/2025 13:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:18
02 Std. Sol.	0.019	13:19
03 Room Air	0.000	13:19
04 Std. Sol.	0.020	13:20
05 Room Air	0.000	13:20
06 Std. Sol.	0.020	13:21
07 Room Air	0.000	13:22

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
04/04/2025 13:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Sol.	0.295	13:25
03 Room Air	0.000	13:25
04 Std. Sol.	0.300	13:26
05 Room Air	0.000	13:26
06 Std. Sol.	0.300	13:27
07 Room Air	0.000	13:28

08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
04/04/2025 13:28

Test	AC	Time
01 Room Air	0.000	13:29
02 *Subject Test	INT*	13:29
03 Room Air	0.000	13:30

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 41
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
04/04/2025 13:35

Test	AC	Time
01 Diagnostic	OK	13:36
02 Room Air	0.000	13:36
03 *Subject Test	RFI*	13:37
04 Room Air	0.000	13:37

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 41
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
04/04/2025 13:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Gas	0.083	13:38
03 Room Air	0.000	13:39
04 Std. Gas	0.082	13:39
05 Room Air	0.000	13:40
06 Std. Gas	0.083	13:40
07 Room Air	0.000	13:40

Lot No = 28423080A3
Cyl No = 41
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3 acn
4/14/25
calibration adjustment needed.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
04/04/2025 13:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Gas	0.083	13:46
03 Room Air	0.000	13:46
04 Std. Gas	0.082	13:47
05 Room Air	0.000	13:47
06 Std. Gas	0.083	13:47
07 Room Air	0.000	13:48

Lot No = 28423080A3
Cyl No = 41
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check - pre cal adj.
calibration adjustment needed.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
06/06/2025 08:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:55
02 Std. Sol.	0.017	08:56
03 Room Air	0.000	08:57
04 Std. Sol.	0.018	08:57
05 Room Air	0.000	08:58
06 Std. Sol.	0.017	08:59
07 Room Air	0.000	08:59


08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC CHECK - 0.020 AC
post calibration adjustment

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
06/06/2025 09:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:02
02 Std. Sol.	0.403	09:03
03 Room Air	0.000	09:04
04 Std. Sol.	0.403	09:04
05 Room Air	0.000	09:05
06 Std. Sol.	0.404	09:06
07 Room Air	0.000	09:06


08 Sim Temp = 34.0°C

Simul Ser No = MP5318

Std Sol No = 20212D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC check -0.400AC
post calibration adjustment

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
06/05/2025 09:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Gas	0.080	09:27
03 Room Air	0.000	09:28
04 Std. Gas	0.080	09:28
05 Room Air	0.000	09:29
06 Std. Gas	0.080	09:29
07 Room Air	0.000	09:30

Lot No = 14323080A4
Cyl No = 41
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3
post calibration adjustment

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
06/05/2025 09:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:33
02 Std. Gas	0.081	09:34
03 Room Air	0.000	09:34
04 Std. Gas	0.080	09:35
05 Room Air	0.000	09:35
06 Std. Gas	0.081	09:36
07 Room Air	0.000	09:36

Lot No = 14323080A4
Cyl No = 41
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #4-6
post calibration adjustment*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = TOXL 8164.14.00 09/16
06/05/2025 09:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Gas	0.080	09:38
03 Room Air	0.000	09:38
04 Std. Gas	0.081	09:39
05 Room Air	0.000	09:39
06 Std. Gas	0.080	09:40
07 Room Air	0.000	09:40

Lot No = 14323080A4
Cyl No = 41
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9
post calibration adjustment.

Form 106-I8000