



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4196

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:

Gas Cylinder Yes or No (If Yes, Lot # 08425080A1 Cyl. # 14)

Keys Yes or No

Power Cord Yes or No

2. ☒ Download Data

3. ☒ Upload Operator File

4. ☒ Current Location Code: HETT

5. ☒ Battery Check

Was the external battery pack replaced? Yes or No

6. ☒ O-Rings

Replaced Simulator O-Ring Yes or No

Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode

2. ☒ Breath tube heated

3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.

Time Zone: CST or MDT (Time on test records will be in time zone circled).

4. ☒ Print test (Level 1,P). Sign and attach test record.

5. ☒ Tank monitor (Level 3,D,G).

Display: 987 psi Regulator: 1000 psi

Display and Regulator $\pm$ 50 psi of each other Yes or No

Gas tank tare necessary? Yes or No

If Yes, display readings after tare (Level 3,M,C,G):

Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - Low AC (Level 1,C)

Known Value $\leq$ 0.03 AC: 0.020 AC

Sim. Ser #: MP5321

Lot #: 202311D

Exp. Date: 11/28/25

☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 2/14/24
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 28423080A3
Cylinder #: 40
Exp. Date: 11/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: 1CS8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.079 Test 4 0.079 Test 7 0.079
Test 2 0.079 Test 5 0.079 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079
☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

18 June 2025
Date

[Signature]
Reviewer

20 June 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 46
Standard Expiration? 11/05/2025
Oper No? 133237

Flow Cal. Date: 04/15/2020
Slope 657
Intercept -683120

IR Calibration Date: 04/16/2020
3um 9um

0th Coef(*100):	-35275	-26830
1st Coef(*100):	271852	134658
2nd Coef(*100):	1067	849
H2O adj(mg/l*10k):	413	302

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 10:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Sol.	0.019	10:57
03 Room Air	0.000	10:58
04 Std. Sol.	0.019	10:59
05 Room Air	0.000	10:59
06 Std. Sol.	0.019	11:00
07 Room Air	0.000	11:01

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Low Ac Check - 0.020Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:02
02 Std. Sol.	0.294	11:03
03 Room Air	0.000	11:04
04 Std. Sol.	0.296	11:05
05 Room Air	0.000	11:05
06 Std. Sol.	0.298	11:06
07 Room Air	0.000	11:06

08 Sim Temp = 34.0°C

Simul Ser No = MP6035
Std Sol No = 202402C
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:08

Test	AC	Time
01 Room Air	0.000	11:08
02 *Subject Test	INT*	11:09
03 Room Air	0.000	11:10

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 46
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent Check*
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:10

Test	AC	Time
01 Diagnostic	OK	11:11
02 Room Air	0.000	11:11
03 *Subject Test	RFI*	11:12
04 Room Air	0.000	11:12

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 46
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:14
02 Std. Gas	0.079	11:14
03 Room Air	0.000	11:15
04 Std. Gas	0.079	11:15
05 Room Air	0.000	11:16
06 Std. Gas	0.079	11:16
07 Room Air	0.000	11:17

Lot No = 28423080A3
Cyl No = 46
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.079	11:18
03 Room Air	0.000	11:18
04 Std. Gas	0.079	11:19
05 Room Air	0.000	11:19
06 Std. Gas	0.079	11:20
07 Room Air	0.000	11:20

Lot No = 28423080A3
Cyl No = 46
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

dry cal check # 4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = TOXL 8164.14.00 09/16
06/18/2025 11:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.079	11:22
03 Room Air	0.000	11:22
04 Std. Gas	0.079	11:22
05 Room Air	0.000	11:23
06 Std. Gas	0.079	11:23
07 Room Air	0.000	11:24

Lot No = 28423080A3
Cyl No = 46
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000