



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3064 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # 01524080A4 Cyl. # 17)  
Keys Yes or No  
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: MCSO
5. ☒ Battery Check  
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).  
Display: 902 psi Regulator: 900 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5321  
Lot #: 202311D  
Exp. Date: 11/28/25  
☒ Results  $\pm$  0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - High AC (Level 1,C)  
Known Value  $\geq 0.25$  AC: 0.400 AC  
Sim. Ser #: MP5318  
Lot #: 2023120  
Exp. Date: 12/7/25  
☒ Results  $\pm 5\%$  AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).  
Known Value: 0.080 AC 6/4/25 new  
Gas Cylinder Lot #: 0.08-14323080A1  
Cylinder #: 72  
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)  
Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7352  
Lot #: ICS8  
Exp. Date: None  
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)  
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
- |                      |                     |                     |
|----------------------|---------------------|---------------------|
| Test 1 <u>0.079</u>  | Test 4 <u>0.079</u> | Test 7 <u>0.080</u> |
| Test 2 <u>0.079</u>  | Test 5 <u>0.079</u> | Test 8 <u>0.080</u> |
| Test 3 <u>0.079</u>  | Test 6 <u>0.079</u> | Test 9 <u>0.079</u> |
| Average <u>0.079</u> |                     |                     |
- ☒ Results  $\pm 0.005$  AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]  
Inspector Signature

04 June 2025  
Date

[Signature]  
Reviewer

04 June 2025  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      13:45

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      01524080A4  
Standard Cyl #?      17  
Standard Expiration?      03/05/2026  
Oper No?      133237

Flow Cal. Date:      07/29/2010  
Slope      806  
Intercept      -895033

IR Calibration Date:      07/29/2010  
   3um      9um

-----  
0th Coef(\*100):      -14075      -5808  
1st Coef(\*100):      272113      138204  
2nd Coef(\*100):      1637      986  
H2O adj(mg/l\*10k): 289      148

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      13:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:46
02 Std. Sol.	0.020	13:47
03 Room Air	0.000	13:48
04 Std. Sol.	0.020	13:48
05 Room Air	0.000	13:49
06 Std. Sol.	0.020	13:50
07 Room Air	0.000	13:50

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: **Low AC check -0.020AC**

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      13:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.393	13:52
03 Room Air	0.000	13:53
04 Std. Sol.	0.393	13:54
05 Room Air	0.000	13:54
06 Std. Sol.	0.392	13:55
07 Room Air	0.000	13:55

08 Sim Temp = 34.0°C

Simul Ser No = MP5318  
Std Sol No = 202312D  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.400AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      13:57

Test	AC	Time
01 Room Air	0.000	13:57
02 *Subject Test	INT*	13:58
03 Room Air	0.000	13:58

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A1  
Cyl No = 72  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Interferent check  
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      13:59

Test	AC	Time
01 Diagnostic	OK	13:59
02 Room Air	0.000	14:00
03 Reference	RFI*	14:00
04 Room Air	0.000	14:00

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A1  
Cyl No = 72  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Gas	0.079	14:01
03 Room Air	0.000	14:02
04 Std. Gas	0.079	14:02
05 Room Air	0.000	14:03
06 Std. Gas	0.079	14:03
07 Room Air	0.000	14:04

Lot No = 14323080A1  
Cyl No = 72  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: DRY CAL CHECK #1-3

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      14:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.079	14:05
03 Room Air	0.000	14:05
04 Std. Gas	0.079	14:06
05 Room Air	0.000	14:06
06 Std. Gas	0.079	14:07
07 Room Air	0.000	14:07

Lot No = 14323080A1  
Cyl No = 72  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.14.00 09/16  
06/04/2025      14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:08
02 Std. Gas	0.080	14:08
03 Room Air	0.000	14:09
04 Std. Gas	0.080	14:09
05 Room Air	0.000	14:10
06 Std. Gas	0.079	14:10
07 Room Air	0.000	14:11

Lot No = 14323080A1  
Cyl No = 72  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000