



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6492 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # 14323080A1 Cyl. # 17)  
Keys Yes or No  
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: FYPD
5. ☒ Battery Check  
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).  
Display: 1000 psi Regulator: 1000 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5321  
Lot #: 202311D  
Exp. Date: 11/28/25  
☒ Results  $\pm$  0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - High AC (Level 1,C)  
Known Value  $\geq 0.25$  AC: 0.400AC AC  
Sim. Ser #: MP5318  
Lot #: 202312D  
Exp. Date: 12/7/25  
☒ Results  $\pm 5\%$  AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).  
Known Value : 0.080 AC  
Gas Cylinder Lot #: 28423080A3  
Cylinder #: 41  
Exp. Date: 11/5/25
4. ☒ Interferent Check (Level 1,B)  
Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7352  
Lot #: ICS 8  
Exp. Date: None  
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)  
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
- |        |              |         |              |        |              |
|--------|--------------|---------|--------------|--------|--------------|
| Test 1 | <u>0.079</u> | Test 4  | <u>0.080</u> | Test 7 | <u>0.080</u> |
| Test 2 | <u>0.080</u> | Test 5  | <u>0.080</u> | Test 8 | <u>0.081</u> |
| Test 3 | <u>0.080</u> | Test 6  | <u>0.080</u> | Test 9 | <u>0.080</u> |
|        |              | Average | <u>0.080</u> |        |              |
- ☒ Results  $\pm 0.005$  AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]  
Inspector Signature

06 June 2025  
Date

[Signature]  
Reviewer

07 June 2025  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      08:45

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      14323080A1  
Standard Cyl #?      17  
Standard Expiration?      06/05/2025  
Oper No?      133237

Flow Cal. Date:      08/18/2015  
Slope      669  
Intercept      -631515

IR Calibration Date:      08/18/2015  
   3um      9um

0th Coef(*100):	-41685	-23145
1st Coef(*100):	269226	142194
2nd Coef(*100):	2671	1239
H2O adj(mg/l*10k):	872	437

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
ANNA NAREHOOD

Remarks:      Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      08:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:47
02 Std. Sol.	0.019	08:47
03 Room Air	0.000	08:48
04 Std. Sol.	0.019	08:49
05 Room Air	0.000	08:49
06 Std. Sol.	0.019	08:50
07 Room Air	0.000	08:50


08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      08:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:52
02 Std. Sol.	0.396	08:53
03 Room Air	0.000	08:54
04 Std. Sol.	0.397	08:55
05 Room Air	0.000	08:55
06 Std. Sol.	0.397	08:56
07 Room Air	0.000	08:57

08 Sim Temp = 34.0°C

Simul Ser No = MP5318

Std Sol No = 202312D

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.400AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      09:01

Test	AC	Time
01 Room Air	0.000	09:02
02 *Subject Test	INT*	09:03
03 Room Air	0.000	09:04

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 41  
Expiration Date = 11/05/2025  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*  
*0.10AC + 0.05% Acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      09:04

Test	AC	Time
01 Diagnostic	OK	09:05
02 Room Air	0.000	09:05
03 *Subject Test	RFI*	09:06
04 Room Air	0.000	09:06

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 41  
Expiration Date = 11/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      09:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:07
02 Std. Gas	0.079	09:07
03 Room Air	0.000	09:08
04 Std. Gas	0.080	09:08
05 Room Air	0.000	09:09
06 Std. Gas	0.080	09:09
07 Room Air	0.000	09:10

Lot No = 28423080A3  
Cyl No = 41  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      09:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:11
02 Std. Gas	0.080	09:11
03 Room Air	0.000	09:12
04 Std. Gas	0.080	09:12
05 Room Air	0.000	09:13
06 Std. Gas	0.080	09:13
07 Room Air	0.000	09:14

Lot No = 28423080A3  
Cyl No = 41  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.14.00 09/16  
06/06/2025      09:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:15
02 Std. Gas	0.080	09:15
03 Room Air	0.000	09:16
04 Std. Gas	0.081	09:16
05 Room Air	0.000	09:17
06 Std. Gas	0.080	09:17
07 Room Air	0.000	09:18

Lot No = 28423080A3  
Cyl No = 41  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #7-9

Form 106-I8000