



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Tyler Molloy		
Location Minot AFB		Intoxilyzer® Serial Number 80-006501
Gas Lot Number 25024080A3	Gas Cylinder Number 50	Gas Expiration Date 10-05-2026

Check When Done:

☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

☒ 2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0.010 AC

B. 0.019 AC

C. 0.079 AC

☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

☒ 5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature 	Date 1 July 2025
Reviewed By (Crime Laboratory Use Only) Anna Narehood	Date 15 July 2025

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = MNAF 8164.14.00 09/16
07/01/2025 15:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.080	15:05
03 Room Air	0.000	15:05
04 Std. Gas	0.079	15:06
05 Room Air	0.000	15:06
06 Std. Gas	0.079	15:07
07 Room Air	0.000	15:07

Lot No = 25024080A3
Cyl No = 50
Exp Date = 10/05/2026
County = 64 Oper No. = 134040


Operator Signature
TYLER MOLLOY

Remarks:

Form 106-I8000