ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print)		
Frank Eberle		
Location: Bowman co. Sheriff's Off.		Intoxilyzer® Serial Number: 80-003071
Gas Lot Number: 28423080 A3	Gas Cylinder Number: 007	Gas Expiration Date: 11/05/2025

Check When Done:

- [x] 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- [x] 2. Perform an ACA Test (Level 1, Function C).

  Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

  A. 0.080 AC
  B. 0.080 AC
  C. 0.080 AC
- [x] 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- [x] 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- [x] 5. Send the Following to the Crime Laboratory:
  A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  B. ACA Test Record.

Chemical Test Operator Signature	Date: 4-24-24
Reviewed By (Crime Laboratory Use Only): Anna Narehood	Date: 05Nov2025

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = BOWM 8164.14.00 09/16
04/24/2024 09:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Gas	0.080	10:01
03 Room Air	0.000	10:01
04 Std. Gas	0.080	10:02
05 Room Air	0.000	10:02
06 Std. Gas	0.080	10:03
07 Room Air	0.000	10:03

Lot No = 28423080A3
Cyl No = 007
Exp Date = 11/05/2025
County = 06 Oper No. = 130015

Operator Signature
FRANK EBERLE

Remarks:

Form 106-I8000