




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number 80-006512	Instrument Location GFCC
Reason for Install/Repair ☐ Install After Receiving From Crime Laboratory ☐ Install After Location Change ☒ Other (Specify) Transported to UND for download/Brought back to GFCC and installed	


Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): **Downloaded at UND**
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  #642	Date 11-15-2025
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen 	Certified Date 20Nov2025
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = GFSO 8164.14.00 09/16
11/18/2025 14:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01524080A4
Standard Cyl #? 18
Standard Expiration? 03/05/2026
Oper No? 133510

Flow Cal. Date: 08/20/2015
Slope 680
Intercept -660856

IR Calibration Date: 08/21/2015
3um 9um

0th Coef(*100):	-23051	-19261
1st Coef(*100):	266836	137135
2nd Coef(*100):	3974	1466
H2O adj(mg/l*10k):	785	487

***** Printer Test End *****



Operator Signature
ADAM COLE

Remarks:

Install

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = GFSO 8164.14.00 09/16
11/18/2025 14:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Gas	0.083	14:36
03 Room Air	0.000	14:37
04 Std. Gas	0.082	14:37
05 Room Air	0.000	14:37
06 Std. Gas	0.083	14:38
07 Room Air	0.000	14:38

Lot No = 01524080A4
Cyl No = 18
Exp Date = 03/05/2026
County = 18 Oper No. = 133510

 #642

Operator Signature
ADAM COLE

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = GFSO 8164.14.00 09/16
11/18/2025 14:40

Test	AC	Time
01 Room Air	RFI*	14:41
02 Room Air	0.000	14:41

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST T
Sub DOB = 01/01/2000
Sub Sex = Male Weight = 150
Test = DUI Cit = 1
Dr. Lic. = ND/NA
Lot No = 01524080A4
Cyl No = 18
Expiration Date = 03/05/2026
County = 18 Oper No. = 133510



Operator Signature
ADAM COLE

Remarks:

Form 106-I8000