



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 REPAIR AND MAINTENANCE

Intoxilyzer® 8000 Serial Number: 80-00 4948 Maintenance Location: TOXL

1. Reported Issue and Issue Found:

During field use, instrument was reported to be
having high calibration values.
low aci 11/28/25

2. Repair/Maintenance Performed:

calibration adjustment required.

3. Does a Calibration Adjustment need to be performed? Yes or No

- If Yes, perform a calibration check.

☒ Configure dry gas standard (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 20

Exp. Date: 9/5/27

☒ Dry Calibration Check (Level 1,C)

Test 1 0.076 Test 2 0.076 Test 3 0.075

- If No, go to step 4.

4. Does the instrument flow sensor need to be adjusted? Yes or No

- If Yes, complete Flow Sensor Adjustment and Calibration (Document ID: 11872), then go to step 5.
- If No, go to step 5.

5. Does the instrument calibration need to be adjusted? Yes or No
- If Yes, complete Intoxilyzer® 8000 Calibration Adjustment (Document ID: 11859), then go to step 6.
 - If No, go to step 6.
6. Was the instrument cover removed for the repair? Yes or No
- If Yes, complete Intoxilyzer® 8000 Initial Inspection (Document ID: 11784), then go to step 7.
 - If No, go to step 7.
7. Is an Annual Inspection due on the Intoxilyzer® 8000? Yes or No
- If Yes, complete Intoxilyzer® 8000 Annual Inspection (Document ID: 11698), step 8 is not needed, and sign below.
 - If No, go to step 8.
8. Was an Intoxilyzer® 8000 Initial Inspection (Document ID: 11784) completed in step 6? Yes or No
- If Yes, no further action is needed and sign below.
 - If No, Complete Intoxilyzer® 8000 Calibration (Document ID: 11871) and sign below.

Remarks/Notes: NIA

Instrument is acceptable to be used in the field.

Yes or No

If No, state reason(s) why:

Alan Engman
Breath Alcohol Analyst Signature

21 Nov 2025
Date

Janette Putschner
Reviewer Signature

26 Nov 2025
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = ~~WCPD ToxL~~ 8164.14.00 09/16
~~09/22/2025~~ **11/21/2025** ~~23:44~~ **13:44**

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:44 13:44
02 Std. Gas	0.076	23:45 13:45
03 Room Air	0.000	23:45 13:45
04 Std. Gas	0.076	23:45 13:45
05 Room Air	0.000	23:46 13:46
06 Std. Gas	0.075	23:46 13:46
07 Room Air	0.000	23:47 13:47

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: **Pre-cal adjustment.**
Incorrect time, date + location
Code.
Form 106-I8000