



ETHANOL GAS STANDARD CYLINDER REPORT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Coby Hubble</i>		
Location <i>Williston Jail</i>	Intoxilyzer® Serial Number <i>80-002669</i>	
Gas Lot Number <i>01524080A4</i>	Gas Cylinder Number <i>24</i>	Gas Expiration Date <i>03/05/2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.081* AC
B. *0.081* AC
C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>J. H.</i>	Date <i>12/18/2025</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i>	Date <i>30Dec2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = WILL 8164.14.00 09/16
12/18/2025 13:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Gas	0.081	13:33
03 Room Air	0.000	13:34
04 Std. Gas	0.081	13:34
05 Room Air	0.000	13:35
06 Std. Gas	0.082	13:35
07 Room Air	0.000	13:36

Lot No = 01524080A4

Cyl No = 24

Exp Date = 03/05/2026

County = 53

Oper No. = 030922

Operator Signature

COBY HUBBLE

Remarks:

Form 106-I8000