



## INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL


CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number	Instrument Location
80-003072	STEE
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change	
<input type="checkbox"/> Other (Specify) _____	


Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature	Date
	<u>6-5-2025</u>

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	Anna Ingemansen		Certified Date
			26Jan2026

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = STEE      8164.14.00 09/16  
06/05/2025      11:00

# \*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNPQRSTUVWXYZ!@#\$%^&\*() +?

## Current Instrument Setup

Data Entry Mode:	Enabled
Start Test Sequence:	DABACABA
Display Prelim Rslt?	Yes
Display Third Digit?	Yes
Inhib Printer(Y/N)?	No
Display Volume?	No
Disable On Memfull?	Yes
# of Print Copies?	1
Select Std (D/W/I)?	Dry
Standard Value?	0.080
Standard Lot #?	01524080A4
Standard Cyl #?	23
Standard Expiration?	03/05/2026
Oper No?	039909

Flow Cal. Date: 03/09/2011  
Slope 682  
Intercept -620512

IR Calibration Date: 03/09/2011  
3um 9um

```

0th Coef(*100) : -24766 -17499
1st Coef(*100) : 264559 138630
2nd Coef(*100) : 1876 1114
H2O adj(mg/l*10k) : 497 331

```

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
BARRY VANNATTA

Remarks: Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = STEE      8164.14.00 09/16  
06/05/2025      11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Gas	0.082	11:01
03 Room Air	0.000	11:02
04 Std. Gas	0.081	11:02
05 Room Air	0.000	11:03
06 Std. Gas	0.081	11:03
07 Room Air	0.000	11:03

Lot No = 01524080A4  
Cyl No = 23  
Exp Date = 03/05/2026  
County = 22      Oper No. = 039909

  
Operator Signature  
BARRY VANNATTA

Remarks: AC A test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = STEE      8164.14.00 09/16  
06/05/2025      11:04

Test	AC	Time
01 Room Air	RFI*	11:06
02 Room Air	RFI*	11:07

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, INSTALL NA  
Sub DOB = 06/09/1987  
Sub Sex = Unknown      Weight = NA  
Test = OTH      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 01524080A4  
Cyl No = 23  
Expiration Date = 03/05/2026  
County = 22      Oper No. = 039909

  
\_\_\_\_\_  
Operator Signature  
BARRY VANNATTA

Remarks: RFI Test

Form 106-I8000