



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Andrew Tunheim</i>		
Location <i>Wishek PD</i>	Intoxilyzer® Serial Number <i>80-005952</i>	
Gas Lot Number <i>19825080A2</i>	Gas Cylinder Number <i>25</i>	Gas Expiration Date <i>9/15/27</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.010* AC
 B. *0.081* AC
 C. *0.010* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>2-2-2026</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i> <i>[Signature]</i>	Date <i>10Feb2026</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = WISH 8164.14.00 09/16
02/02/2026 10:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:51
02 Std. Gas	0.080	10:51
03 Room Air	0.000	10:52
04 Std. Gas	0.081	10:52
05 Room Air	0.000	10:53
06 Std. Gas	0.080	10:53
07 Room Air	0.000	10:54

Lot No = 19825080A2
Cyl No = 25
Exp Date = 09/05/2027
County = 26 Oper No. = 010912



Operator Signature
ANDREW TUNHEIM

Remarks: