



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Noah Pedersen		
Location LaMoure County SO		Intoxilyzer® Serial Number 80-006507
Gas Lot Number 19825080A2	Gas Cylinder Number 19	Gas Expiration Date 9/5/2027

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.080 AC
  - B. 0.080 AC
  - C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature 	Date 2-17-26
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen	Date 17Feb2026

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006507  
Location = LAMO      8164.14.00 09/16  
02/17/2026      10:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:17
02 Std. Gas	0.080	10:17
03 Room Air	0.000	10:18
04 Std. Gas	0.080	10:18
05 Room Air	0.000	10:19
06 Std. Gas	0.080	10:19
07 Room Air	0.000	10:19

Lot No = 19825080A2  
Cyl No = 19  
Exp Date = 09/05/2027  
County = 23      Oper No. = 134259

  
Operator Signature  
NOAH PEDERSEN

Remarks:

Form 106-I8000

The undersigned, having the custody of the original public record, certifies that the information herein is a true and correct copy of the original document maintained as a part of the activities of this office.

Signature / Title   
Office / Agency LCSO