

State of North Dakota)
)ss
County of Burleigh)

I, Janelle Portscheller, do hereby certify that I am the duly-appointed State Toxicologist for the State of North Dakota and an official custodian of the records and files of the office thereof, that I have carefully compared the

Intoximeters Alco-Sensor FST® Ethanol Breath Standard Analytical Report, Lot No. AG533601, Expiration 12/02/27 (02/04/26)

hereto attached with the respective original as the same appears of record on file in the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, and find the same to be a true and correct copy thereof and of the whole thereof. In witness whereof I have set my hand at the city of Bismarck, in said county this:

4th day of February, 2026

Janelle Portscheller
Janelle Portscheller, State Toxicologist


State of North Dakota)
)ss
County of Burleigh)

On this 4th day of February, 2026, before me personally appeared Janelle Portscheller, known to me to be the State Toxicologist for the State of North Dakota, and acknowledged to me that she has executed the same.

Subscribed to and sworn before me this:

4th day of February, 2026

Deanna Dailey
Deanna Dailey, Notary Public, State of North Dakota
My Commission Expires March 23, 2027



(SEAL)

**INTOXIMETERS ALCO-SENSOR FST®
ETHANOL BREATH STANDARD ANALYTICAL REPORT**

Ethanol Breath Standard Lot Number AG533601 Expiration Date 12/02/2027

This standard was analyzed by AIRGAS USA LLC (LAB) with a reported result of 223 ppm which is the equivalent of 0.082 BrAC of Ethanol in Nitrogen. AIRGAS USA LLC (LAB) has provided a Certificate of Analysis traceable to N.I.S.T. RGM Ethanol Standards.

The number of cylinders sent to each location will be based on need. The standard may be used until the date of expiration as indicated by the manufacturer's Certificate of Analysis.



Janelle Portscheller, State Toxicologist



Date Approved



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 2-Dec-2025

Lot # AG533601 Model 108

Exp Date 2-Dec-2027	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.082 ± 0.002 BrAC (223 ppm)
-------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 12.04.2025 18:10



Approved for Release: _____
Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07