



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) CHRISTOPHER ALMLIE		
Location Burke CO SO		Intoxilyzer® Serial Number 80-004947
Gas Lot Number 25024080A3	Gas Cylinder Number 57	Gas Expiration Date 10-05-2026

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.081 AC
 B. 0.082 AC
 C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

- AEI 5. Send the Following to the Crime Laboratory:
 04Mar26 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature 	Date 2/22/26
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen	Date 04Mar2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = BURK 8164.14.00 09/16
02/22/2026 14:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Gas	0.081	14:36
03 Room Air	0.000	14:37
04 Std. Gas	0.082	14:37
05 Room Air	0.000	14:37
06 Std. Gas	0.082	14:38
07 Room Air	0.000	14:38

Lot No = 25024080A3
Cyl No = 57
Exp Date = 10/05/2026
County = 07 Oper No. = 132664



Operator Signature
CHRISTOPHER ALMLIE

Remarks: *good*

Form 106-I8000