



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-004947	Instrument Location Burk Co SD
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 2/22/26
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen	Certified Date 04Mar2026
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = BURK 8164.14.00 09/16
02/22/2026 14:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 25024080A3
Standard Cyl #? 57
Standard Expiration? 10/05/2026
Oper No? 132664

Flow Cal. Date: 04/07/2020
Slope 618
Intercept -565524

IR Calibration Date: 04/09/2024
 3um 9um

0th Coef(*100):	-21002	-23184
1st Coef(*100):	264237	133599
2nd Coef(*100):	1927	1247
H2O adj(mg/l*10k):	409	438

***** Printer Test End *****


Operator Signature
CHRISTOPHER ALMLIE

Remarks:

good

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = BURK 8164.14.00 09/16
02/22/2026 14:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Gas	0.081	14:36
03 Room Air	0.000	14:37
04 Std. Gas	0.082	14:37
05 Room Air	0.000	14:37
06 Std. Gas	0.082	14:38
07 Room Air	0.000	14:38

Lot No = 25024080A3
Cyl No = 57
Exp Date = 10/05/2026
County = 07 Oper No. = 132664



Operator Signature
CHRISTOPHER ALMLIE

Remarks: *good*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = BURK 8164.14.00 09/16
02/22/2026 14:39

Test	AC	Time
01 Room Air	RFI*	14:41
02 Room Air	RFI*	14:41

*Invalid Test
Inhibited - RFI

Sub Name = ALMLIE, CHRISTOPHER LEE
Sub DOB = 03/29/1991
Sub Sex = Male Weight = 260
Test = DUI Cit = NA
Dr. Lic. = ND/ALM910377
Lot No = 25024080A3
Cyl No = 57
Expiration Date = 10/05/2026
County = 07 Oper No. = 132664



Operator Signature
CHRISTOPHER ALMLIE

Remarks: *good*

Form 106-I8000