



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <i>80-005357</i>	Instrument Location <i>GRAF</i>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <i>Checking Modern lines due to it not connecting</i>	

Check When Done:

- 1.- Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>David Ribbo</i>	Date <i>2/26/2026</i>
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Anna Ingemansen</i>	Certified Date <i>05Mar2026</i>
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = GRAF 8164.14.00 09/16
02/26/2026 14:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Gas	0.082	14:12
03 Room Air	0.000	14:13
04 Std. Gas	0.081	14:13
05 Room Air	0.000	14:14
06 Std. Gas	0.082	14:14
07 Room Air	0.000	14:15

Lot No = 19825080A2
Cyl No = 26
Exp Date = 09/05/2027
County = 50

Oper No. = 132597

Jacob Nikko 4714

Operator Signature
JACOB NIKKO

Remarks:

Form 106-I8000

The Undersigned, having custody of the original public record, certifies that the information hereon is a true and correct copy of the original document maintained as a part of the activities of this agency.

Signature/Title *Jacob Nikko* Deputy
Agency/Date *Walsh 50 Co/2/26/26*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = GRAF 8164.14.00 09/16
02/26/2026 14:17


DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:18
02 Std. Gas	RFI*	14:18
03 Room Air	0.000	14:19

*Invalid Test
Inhibited - RFI

Lot No = 19825080A2
Cyl No = 26
Exp Date = 09/05/2027
County = 50

Oper No. = 132597



Operator Signature
JACOB NIKKO

Remarks:

Form 106-I8000

The Undersigned, having custody of the original public record, certifies that the information hereon is a true and correct copy of the original document maintained as a part of the activities of this agency.

Signature/Title Jacob Nikko / Deputy
Agency/Date Walsh Co So / 2/26/26