ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Cole, Adam, W.		
Location GFSO(GFCCC)	Intoxilyzer® Serial Number 80-006512	
Gas Lot Number 264437	Gas Cylinder Number 120	Gas Expiration Date 01-29-2028

Check When Done:

- [x] 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- [x] 2. Perform an ACA Test (Level 1, Function C).

  Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

  A. 0.081 AC
  B. 0.082 AC
  C. 0.081 AC

- [x] 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- [x] 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- [x] 5. Send the Following to the Crime Laboratory:

  A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  B. ACA Test Record.

Chemical Test Operator Signature #642	Date 03-05-2026
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen Ann Ingemansen	Date 09Mar2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer
North Dakota Model 8000
Location = GFSO
03/05/2026

Alcohol Analyzer
SN 80-006512
8164.14.00 09/16
15:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:15
02 Std. Gas	0.081	15:15
03 Room Air	0.000	15:16
04 Std. Gas	0.082	15:16
05 Room Air	0.000	15:17
06 Std. Gas	0.081	15:17
07 Room Air	0.000	15:18

Lot No = 264437
Cyl No = 120
Exp Date = 01/29/2028
County = 18

Oper No. = 133510

#642

Operator Signature
ADAM COLE

Remarks:

Form 106-I8000