



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Beth Isaak</i>		
Location <i>DCSO</i>	Intoxilyzer® Serial Number <i>80-006498</i>	
Gas Lot Number <i>19825080A2</i>	Gas Cylinder Number <i>14</i>	Gas Expiration Date <i>9-5-27</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.*080* AC
 B. 0.*080* AC
 C. 0.*081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>Beth Isaak</i>	Date <i>3-6-26</i>
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen <i>Anna Ingemansen</i>	Date 09Mar2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = DUNN 8164.14.00 09/16
03/06/2026 14:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:21
02 Std. Gas	0.080	14:21
03 Room Air	0.000	14:22
04 Std. Gas	0.080	14:22
05 Room Air	0.000	14:23
06 Std. Gas	0.081	14:23
07 Room Air	0.000	14:24

Lot No = 19825080A2
Cyl No = 14
Exp Date = 09/05/2027
County = 13 Oper No. = 131919



Operator Signature
BETH ISAAK

Remarks:

Form 106-I8000