



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Allen, Kathryn</i>		
Location <i>Cass County Jail - Booking</i>	Intoxilyzer® Serial Number <i>80-005951</i>	
Gas Lot Number <i>19825080A2</i>	Gas Cylinder Number <i>85</i>	Gas Expiration Date <i>09/05/2027</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.079* AC
 - B. *0.079* AC
 - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn M. Allen</i>	Date <i>3-12-26</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i> Anna Ingemansen	Date <i>25Mar2026</i> <i>Anna Ingemansen</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = CASS 8164.14.00 09/16
03/12/2026 17:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:08
02 Std. Gas	0.079	17:08
03 Room Air	0.000	17:09
04 Std. Gas	0.079	17:09
05 Room Air	0.000	17:10
06 Std. Gas	0.079	17:10
07 Room Air	0.000	17:11

Lot No = 19825080A2
Cyl No = 35
Exp Date = 09/05/2027
County = 09 Oper No. = 090801


Operator Signature
KATHRYN ALLEN

Remarks:

All within Normal Ranges!

Form 106-I8000

KAZ677