

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3058 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 19825080 A 2 Cyl. # 24)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: BEPD
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 998 psi Regulator: 1000 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.025 AC
Sim. Ser #: MP5290
Lot #: 202509A
Exp. Date: 9/10/27
 Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP5321

Lot #: 202408H

Exp. Date: 29 Aug 2026

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 264437

Cylinder #: 104

Exp. Date: 1/29/2028

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS 9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.079 Test 4 0.081 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.081

Test 3 0.080 Test 6 0.080 Test 9 0.079

Average 0.080

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

Ann Lyngren
Inspector Signature

01 April 2026
Date

Janelle Putschler
Reviewer

01 Apr 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 11:48

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.800
Standard Lot #? 19825080A2
Standard Cyl #? 24
Standard Expiration? 09/05/2027
Oper No? 133237

Flow Cal. Date: 04/07/2008
Slope 698
Intercept -678970

IR Calibration Date: 04/15/2015
 3um 9um

0th Coef(*100): -18934 -10185
1st Coef(*100): 272557 131040
2nd Coef(*100): 2331 1317
H2O adj(mg/l*10k): 333 280

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 12:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:11
02 Std. Sol.	0.026	12:12
03 Room Air	0.000	12:13
04 Std. Sol.	0.025	12:13
05 Room Air	0.000	12:14
06 Std. Sol.	0.026	12:15
07 Room Air	0.000	12:15


08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 202509A

County = 08

Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *LOW AC check*

0.025AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 12:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:23
02 Std. Sol.	0.296	12:24
03 Room Air	0.000	12:25
04 Std. Sol.	0.296	12:25
05 Room Air	0.000	12:26
06 Std. Sol.	0.296	12:27
07 Room Air	0.000	12:27

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202408H
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: High AC Check
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 12:34

Test	AC	Time
01 Room Air	0.000	12:35
02 Invalid Sample	X.XXX	12:35
03 Room Air	0.000	12:36

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID SAMPLE
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *invalid sample - mouth alcohol*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 12:36

Test	AC	Time
01 Diagnostic	OK	12:37
02 Room Air	RFI*	12:37
03 Room Air	0.000	12:38

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 12:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:39
02 Std. Gas	0.079	12:39
03 Room Air	0.000	12:39
04 Std. Gas	0.080	12:40
05 Room Air	0.000	12:40
06 Std. Gas	0.080	12:41
07 Room Air	0.000	12:41

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.14.00 09/16
04/01/2026 12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:42
02 Std. Gas	0.081	12:42
03 Room Air	0.000	12:43
04 Std. Gas	0.080	12:43
05 Room Air	0.000	12:44
06 Std. Gas	0.080	12:44
07 Room Air	0.000	12:45

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#4-b

Form 106-I8000

