

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 0500 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 01524080A4 Cyl. # 25)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: HARV
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 818 psi Regulator: 800 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.025 AC
Sim. Ser #: MP5290
Lot #: 202509A
Exp. Date: 10 Sept 27
 Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP5321

Lot #: 202408H

Exp. Date: 29 Aug 26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 20

Exp. Date: 9/5/2027

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS 9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.080 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.081

Test 3 0.079 Test 6 0.080 Test 9 0.080

Average 0.080

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

Alex Lyngner
Inspector Signature

02 April 2026
Date

Janelle Putschillo
Reviewer

04 Apr. 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01524080A4
Standard Cyl #? 25
Standard Expiration? 03/05/2026
Oper No? 133237

Flow Cal. Date: 08/18/2015
Slope 681
Intercept -680985

IR Calibration Date: 08/18/2015

	3um	9um
0th Coef(*100):	-3148	-17259
1st Coef(*100):	254475	138520
2nd Coef(*100):	1954	1329
H2O adj(mg/l*10k):	193	414

***** Printer Test End *****


Operator Signature
ANNA INGEMANSEN

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Sol.	0.026	11:11
03 Room Air	0.000	11:12
04 Std. Sol.	0.026	11:13
05 Room Air	0.000	11:13
06 Std. Sol.	0.025	11:14
07 Room Air	0.000	11:14

08 Sim Temp = 34.0°C

Simul Ser No = MP5290
Std Sol No = 202509A
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: LOW AC check
0.025AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:16
02 Std. Sol.	0.300	11:17
03 Room Air	0.000	11:17
04 Std. Sol.	0.301	11:18
05 Room Air	0.000	11:19
06 Std. Sol.	0.301	11:19
07 Room Air	0.000	11:20

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: High AC check
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:21

Test	AC	Time
01 Room Air	0.000	11:32
02 *Subject Test	INT*	11:33
03 Room Air	0.000	11:33

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 20
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent Check
0.10AL + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:35

Test	AC	Time
01 Room Air	0.000	11:36
02 Invalid Sample	X.XXX	11:36
03 Room Air	0.000	11:37

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID SAMPLE
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 20
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Invalid Sample Mouth Alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:37

Test	AC	Time
01 Diagnostic	OK	11:38
02 Room Air	0.000	11:38
03 *Subject Test	RFI*	11:39
04 Room Air	0.000	11:39

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 20
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:40
02 Std. Gas	0.080	11:41
03 Room Air	0.000	11:41
04 Std. Gas	0.080	11:41
05 Room Air	0.000	11:42
06 Std. Gas	0.079	11:42
07 Room Air	0.000	11:43

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *Dry cal check
#1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:44
02 Std. Gas	0.080	11:44
03 Room Air	0.000	11:45
04 Std. Gas	0.080	11:45
05 Room Air	0.000	11:46
06 Std. Gas	0.080	11:46
07 Room Air	0.000	11:47

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/02/2026 11:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:48
02 Std. Gas	0.080	11:49
03 Room Air	0.000	11:49
04 Std. Gas	0.081	11:50
05 Room Air	0.000	11:50
06 Std. Gas	0.080	11:50
07 Room Air	0.000	11:51

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#7-9

Form 106-I8000