



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4956 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 08425080A1 Cyl. # 5)
  - Keys  Yes or No
  - Power Cord  Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: COOP
5.  Battery Check
  - Was the external battery pack replaced?  Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 952 psi Regulator: 950 psi  
Display and Regulator ± 50 psi of each other  Yes or No  
Gas tank tare necessary? Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.025 AC  
Sim. Ser #: MP5290  
Lot #: 202509A  
Exp. Date: 10 Sept. 27  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP5321

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 26-4437

Cylinder #: 104

Exp. Date: 1/29/28

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1CS9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.081 Test 4 0.082 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

*Au Lyngren*  
Inspector Signature

02 Apr 2026  
Date

*Janelle Putscheller*  
Reviewer

04 Apr. 2026  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      12:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Sol.	0.027	12:41
03 Room Air	0.000	12:41
04 Std. Sol.	0.028	12:42
05 Room Air	0.000	12:42
06 Std. Sol.	0.027	12:43
07 Room Air	0.000	12:44

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 202509A

County = 08

Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

LOW AC Check  
0.020 ali 2 Apr 26 0.025 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      12:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Sol.	0.303	12:46
03 Room Air	0.000	12:46
04 Std. Sol.	0.302	12:47
05 Room Air	0.000	12:48
06 Std. Sol.	0.303	12:48
07 Room Air	0.000	12:49

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202408H

County = 08

Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: High AC Check  
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      12:54

Test	AC	Time
01 Room Air	0.000	12:54
02 *Subject Test	INT*	12:55
03 Room Air	0.000	12:56

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 104  
Expiration Date = 01/29/2028  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: Interferent Check  
0.10AC + 0.05% Acetone

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      12:56

Test	AC	Time
01 Room Air	0.000	12:57
02 Invalid Sample	X.XXX	12:57
03 Room Air	0.000	12:58

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INVALID SAMPLE  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 104  
Expiration Date = 01/29/2028  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Invalid Sample  
Mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      12:58

Test	AC	Time
01 Diagnostic	OK	12:59
02 Room Air	RFI*	13:00
03 Room Air	0.000	13:00

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 104  
Expiration Date = 01/29/2028  
County = 08      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      13:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:01
02 Std. Gas	0.081	13:01
03 Room Air	0.000	13:02
04 Std. Gas	0.080	13:02
05 Room Air	0.000	13:02
06 Std. Gas	0.080	13:03
07 Room Air	0.000	13:03

Lot No = 264437  
Cyl No = 104  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *Dry cal check*  
*#1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      13:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:04
02 Std. Gas	0.082	13:05
03 Room Air	0.000	13:05
04 Std. Gas	0.080	13:05
05 Room Air	0.000	13:06
06 Std. Gas	0.080	13:06
07 Room Air	0.000	13:07

Lot No = 264437  
Cyl No = 104  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry Cal check  
#4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = TOXL      8164.14.00 09/16  
04/02/2026      13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:08
02 Std. Gas	0.080	13:08
03 Room Air	0.000	13:09
04 Std. Gas	0.080	13:09
05 Room Air	0.000	13:10
06 Std. Gas	0.080	13:10
07 Room Air	0.000	13:10

Lot No = 264437  
Cyl No = 104  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
# 7-9

Form 106-I8000