



**NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION**

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6509 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 26-4437 Cyl. # 106)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: GFAF
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
 - Display: 753 psi Regulator: 750 psi
 - Display and Regulator ± 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value ≤ 0.03 AC: 0.025 AC
 - Sim. Ser #: MP5290
 - Lot #: 202509A
 - Exp. Date: 9/10/27
 - Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP5321

Lot #: 202408H

Exp. Date: 29 Aug 26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 20

Exp. Date: 9/5/27

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1CS9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.078 Test 4 0.079 Test 7 0.078

Test 2 0.079 Test 5 0.079 Test 8 0.078

Test 3 0.078 Test 6 0.079 Test 9 0.079

Average 0.078

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

Am Lynnen
Inspector Signature

02 April 2026
Date

Janelle Putschiller
Reviewer

04 Apr 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 106
Standard Expiration? 01/29/2028
Oper No? 133237

Flow Cal. Date: 06/11/2019
Slope 445
Intercept 431160

IR Calibration Date: 06/11/2019
 3um 9um

0th Coef(*100): -27266 -14145
1st Coef(*100): 276865 139368
2nd Coef(*100): 2326 1158
H2O adj(mg/l*10k): 584 288

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:16
02 Std. Sol.	0.026	14:17
03 Room Air	0.000	14:18
04 Std. Sol.	0.026	14:18
05 Room Air	0.000	14:19
06 Std. Sol.	0.026	14:20
07 Room Air	0.000	14:20

08 Sim Temp = 34.0°C

Simul Ser No = MP5290
Std Sol No = 202509A
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *LOW AC check*
0.025Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Sol.	0.296	14:23
03 Room Air	0.000	14:23
04 Std. Sol.	0.298	14:24
05 Room Air	0.000	14:24
06 Std. Sol.	0.297	14:25
07 Room Air	0.000	14:26

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: High AC check
0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:27

Test	AC	Time
01 Room Air	0.000	14:28
02 *Subject Test	INT*	14:28
03 Room Air	0.000	14:29

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 20
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent check

0.10AC + 0.05% Acetone

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:29

Test	AC	Time
01 Room Air	0.000	14:30
02 Invalid Sample	X.XXX	14:30
03 Room Air	0.000	14:31

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID SAMPLE
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 20
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Invalid test
 Mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:31

Test	AC	Time
01 Diagnostic	OK	14:32
02 Room Air	0.000	14:32
03 *Subject Test	RFI*	14:33
04 Room Air	0.000	14:34

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 20
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:35
02 Std. Gas	0.078	14:35
03 Room Air	0.000	14:36
04 Std. Gas	0.079	14:36
05 Room Air	0.000	14:36
06 Std. Gas	0.078	14:37
07 Room Air	0.000	14:37

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Gas	0.079	14:41
03 Room Air	0.000	14:42
04 Std. Gas	0.079	14:42
05 Room Air	0.000	14:43
06 Std. Gas	0.079	14:43
07 Room Air	0.000	14:44

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *Dry cal check*
4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
04/02/2026 14:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.078	14:45
03 Room Air	0.000	14:45
04 Std. Gas	0.078	14:46
05 Room Air	0.000	14:46
06 Std. Gas	0.079	14:47
07 Room Air	0.000	14:47

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *Dry Cal check
#7-9*

Form 106-I8000