



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|---|--|
| Serial Number 80-003058 | Instrument Location BEULAH POLICE DEPARTMENT |
| Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____ | |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | |
|--|-----------------------|
| Field Inspector Signature [Signature] 039714 | Date 4-7-26 |
|--|-----------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|--|--------------------------------------|
| Reviewed/Certified By Anna Ingemansen <i>Anna Ingemansen</i> | Certified Date 07April2026 |
|--|--------------------------------------|

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = BEPD 8164.14.00 09/16
04/02/2026 13:48

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &* () _ + ?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &* () _ + ?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19825080A2
Standard Cyl #? 24
Standard Expiration? 09/05/2027
Oper No? 039714

Flow Cal. Date: 04/07/2008
Slope 698
Intercept -678970

IR Calibration Date: 04/15/2015
 3um 9um

0th Coef (*100): -18934 -10185
1st Coef (*100): 272557 131040
2nd Coef (*100): 2331 1317
H2O adj (mg/l*10k): 333 280

***** Printer Test End *****

 039714

Operator Signature
FRANK SENN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = BEPD 8164.14.00 09/16
04/02/2026 14:10

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:10 |
| 02 Std. Gas | 0.079 | 14:11 |
| 03 Room Air | 0.000 | 14:11 |
| 04 Std. Gas | 0.079 | 14:12 |
| 05 Room Air | 0.000 | 14:12 |
| 06 Std. Gas | 0.079 | 14:13 |
| 07 Room Air | 0.000 | 14:13 |

Lot No = 19825080A2
Cyl No = 24
Exp Date = 09/05/2027
County = 29 Oper No. = 039714



Operator Signature
FRANK SENN

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = BEPD 8164.14.00 09/16
04/02/2026 14:14

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 14:15 |
| 02 *Subject Test | RFI* | 14:16 |
| 03 Room Air | 0.000 | 14:17 |

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 09/09/2009
Sub Sex = Male Weight = 125
Test = DUI Cit = 0000000000
Dr. Lic. = ND/TES909111
Lot No = 19825080A2
Cyl No = 24
Expiration Date = 09/05/2027
County = 29 Oper No. = 039714


Operator Signature
FRANK SENN

Remarks:

Form 106-I8000