



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Hella Norby</i>		
Location <i>HARV</i>	Intoxilyzer® Serial Number <i>80-006500</i>	
Gas Lot Number <i>26-4437</i>	Gas Cylinder Number <i>125</i>	Gas Expiration Date <i>01/29/2028</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.080* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Hella Norby</i>	Date <i>04/03/2026</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i>	Date <i>09April2026</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.14.00 09/16
04/03/2026 10:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:41
02 Std. Gas	0.082	10:42
03 Room Air	0.000	10:42
04 Std. Gas	0.080	10:43
05 Room Air	0.000	10:43
06 Std. Gas	0.081	10:44
07 Room Air	0.000	10:44

Lot No = 26 4437
Cyl No = 125
Exp Date = 01/29/2028
County = 52 Oper No. = 133723



Operator Signature

KELLA NORBY

Remarks:

ACA Test

Form 106-I8000