



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-003060</b>	Instrument Location <b>WCPD</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P). AEI 21April2026. RFI test was attached.
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>4-3-26</b>
-------------------------------	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Anna Ingemansen</b> 	Certified Date <b>21April2026</b>
---	--------------------------------------




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = WCPD      8164.14.00 09/16  
04/03/2026      18:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:32
02 Std. Gas	0.080	18:32
03 Room Air	0.000	18:33
04 Std. Gas	0.080	18:33
05 Room Air	0.000	18:33
06 Std. Gas	0.080	18:34
07 Room Air	0.000	18:34

Lot No = 19825080A2  
Cyl No = 27  
Exp Date = 09/05/2027  
County = 27      Oper No. = 131748

  
Operator Signature  
JONATHAN DAVIS

Remarks: *instal*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = WCPD      8164.14.00 09/16  
04/03/2026      18:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	18:35
02 Room Air	0.000	18:36

\*Invalid Test  
Inhibited - RFI

Lot No = 19825080A2  
Cyl No = 27  
Exp Date = 09/05/2027  
County = 27      Oper No. = 131748



Operator Signature  
JONATHAN DAVIS

Remarks:

*inst 4*

Form 106-I8000