



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Michael Dehn</i>		
Location <i>CROS</i>	Intoxilyzer® Serial Number <i>80-006682</i>	
Gas Lot Number <i>26-4437</i>	Gas Cylinder Number <i>121</i>	Gas Expiration Date <i>1/29/28</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0. 079 AC
 B. 0. 079 AC
 C. 0. 079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Michael Dehn</i>	Date <i>04/06/26</i>
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen <i>Anna Ingemansen</i>	Date 22April2026

The undersigned, having custody of the original record, certifies that the information hereon is a true and correct copy of the original document maintained as part of the activities of this office.

Michael Dehn *04/06/26*
 Signature Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = CROS 8164.14.00 09/16
04/06/2026 19:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:32
02 Std. Gas	0.079	19:32
03 Room Air	0.000	19:33
04 Std. Gas	0.079	19:33
05 Room Air	0.000	19:33
06 Std. Gas	0.079	19:34
07 Room Air	0.000	19:34


Lot No = 264437

Cyl No = 121

Exp Date = 01/29/2028

County = 12

Oper No. = 130831



Operator Signature
MICHAEL DEHN

Remarks:

Form 106-I8000