



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <u>80-004942</u>	Instrument Location <u>FAPD</u>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Reinstallation After Annual Inspection at Crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

AEI
22April2026

Field Inspector Signature <u>[Signature]</u>	Date <u>4/6/2026</u>
---	-------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <u>Anna Ingemansen</u>	<u>[Signature]</u>	Certified Date <u>22April2026</u>
---	--------------------	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = FAPD 8164.14.00 09/16
04/06/2026 18:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19825080A2
Standard Cyl #? 033
Standard Expiration? 09/05/2027
Oper No? 130056

Flow Cal. Date: 05/25/2018
Slope 626
Intercept -548004

IR Calibration Date: 04/09/2024

	3um	9um
0th Coef(*100):	-22609	-23575
1st Coef(*100):	260445	132718
2nd Coef(*100):	2885	1338
H2O adj(mg/l*10k):	492	488

***** Printer Test End *****



Operator Signature
PATRICK THOMAS

Remarks: *Good test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = FAPD 8164.14.00 09/16
04/06/2026 18:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:33
02 Std. Gas	0.080	18:34
03 Room Air	0.000	18:34
04 Std. Gas	0.080	18:35
05 Room Air	0.000	18:35
06 Std. Gas	0.080	18:35
07 Room Air	0.000	18:36

Lot No = 19825080A2
Cyl No = 033
Exp Date = 09/05/2027
County = 09 Oper No. = 130056



Operator Signature
PATRICK THOMAS

Remarks: *Valid test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = FAPD 8164.14.00 09/16
04/06/2026 18:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	18:44
02 Room Air	0.000	18:44

*Invalid Test
Inhibited - RFI

Lot No = 19825080A2
Cyl No = 033
Exp Date = 09/05/2027
County = 09 Oper No. = 130056


Operator Signature
PATRICK THOMAS

Remarks:

Form 106-I8000