



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-006674</b>	Instrument Location <b>Washburn</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>4/6/26</b>
-------------------------------	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Anna Ingemansen</b> 	Certified Date <b>22April2026</b>
---	--------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006674  
Location = WASH                8164.14.00 09/16  
04/06/2026                      14:22

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                08425080A1  
Standard Cyl #?                06  
Standard Expiration?           05/05/2027  
Oper No?                        100724

Flow Cal. Date:                06/28/2017  
Slope                            707  
Intercept                       -708478

IR Calibration Date:            06/28/2017  
                                  3um                            9um

-----  
0th Coef (\*100):            -3915                        -18046  
1st Coef (\*100):            256520                      127506  
2nd Coef (\*100):            3843                        1644  
H2O adj (mg/l\*10k):        428                         568

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CODY MEADOWS

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006674  
Location = WASH      8164.14.00 09/16  
04/06/2026      14:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:59
02 Std. Gas	0.081	15:00
03 Room Air	0.000	15:00
04 Std. Gas	0.081	15:01
05 Room Air	0.000	15:01
06 Std. Gas	0.082	15:01
07 Room Air	0.000	15:02

Lot No = 08425080A1  
Cyl No = 06  
Exp Date = 05/05/2027  
County = 28      Oper No. = 100724



Operator Signature  
CODY MEADOWS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006674  
Location = WASH                8164.14.00 09/16  
04/06/2026                        15:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	15:04
02 Room Air	RFI*	15:04

\*Invalid Test  
Inhibited - RFI

Lot No = 08425080A1  
Cyl No = 06  
Exp Date = 05/05/2027  
County = 28                      Oper No. = 100724



Operator Signature  
CODY MEADOWS

Remarks:

Form 106-I8000