



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Trevor Bergerson</i>		
Location <i>BOPD</i>	Intoxilyzer® Serial Number <i>80-007093</i>	
Gas Lot Number <i>26-4437</i>	Gas Cylinder Number <i>103</i>	Gas Expiration Date <i>1/29/2028</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0. *080* AC
  - B. 0. *080* AC
  - C. 0. *079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>T. Bergerson</i> #1821	Date <i>4/7/2026</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i>	Date <i>23April2026</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007093  
Location = BOPD      8164.16.00 09/18  
04/07/2026      17:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:59
02 Std. Gas	0.080	17:59
03 Room Air	0.000	18:00
04 Std. Gas	0.080	18:00
05 Room Air	0.000	18:01
06 Std. Gas	0.079	18:01
07 Room Air	0.000	18:02

Lot No = 264437  
Cyl No = 103  
Exp Date = 01/29/2028  
County = 06      Oper No. = 130538

  
Operator Signature  
TREVOR BERGERSON

Remarks:

Install.

Form 106-I8000