



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3061 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 25024080A3 Cyl. # 8)
  - Keys  Yes or  No
  - Power Cord  Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: LANG
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
  - Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).
  - Display: 999 psi Regulator: 1000 psi
  - Display and Regulator ± 50 psi of each other  Yes or  No
  - Gas tank tare necessary? Yes or  No
  - If Yes, display readings after tare (Level 3,M,C,G):
  - Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).
  - Wet Calibration Check - Low AC (Level 1,C)
  - Known Value ≤ 0.03 AC: 0.025 AC
  - Sim. Ser #: MP3063
  - Lot #: 202509A
  - Exp. Date: 9/10/27
  - Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 26-4437

Cylinder #: 104

Exp. Date: 1/29/28

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.079 Test 4 0.079 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.079

Test 3 0.080 Test 6 0.079 Test 9 0.080

Average 0.079

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. (Yes) or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

*Aw Lyng*  
Inspector Signature

09 April 2026  
Date

*Jocelle Putschler*  
Reviewer

10 Apr. 2026  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      13:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Sol.	0.024	13:38
03 Room Air	0.000	13:39
04 Std. Sol.	0.023	13:40
05 Room Air	0.000	13:40
06 Std. Sol.	0.024	13:41
07 Room Air	0.000	13:41

08 Sim Temp = 34.0°C

Simul Ser No = MP3063  
Std Sol No = 202509A  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *Low AC Check*  
*0.025AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      13:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Sol.	0.297	13:44
03 Room Air	0.000	13:44
04 Std. Sol.	0.299	13:45
05 Room Air	0.000	13:46
06 Std. Sol.	0.299	13:46
07 Room Air	0.000	13:47


08 Sim Temp = 34.0°C

Simul Ser No = MP3003

Std Sol No = 202408H

County = 08

Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

High AC Check  
0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      13:48

Test	AC	Time
01 Room Air	0.000	13:48
02 *Subject Test	INT*	13:49
03 Room Air	0.000	13:50

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 104  
Expiration Date = 01/29/2028  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *Interferent check*  
*0.10AC + 0.05% Acetone*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      13:53

Test	AC	Time
01 Room Air	0.000	13:54
02 Invalid Sample	X.XXX	13:54
03 Room Air	0.000	13:55

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 104  
Expiration Date = 01/29/2028  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *Invalid sample  
mouth alcohol*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      14:00

Test	AC	Time
01 Diagnostic	OK	14:01
02 Room Air	RFI*	14:01
03 Room Air	0.000	14:02

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 104  
Expiration Date = 01/29/2028  
County = 08      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA INGEMANSEN

Remarks: *RFI check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      14:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Gas	0.079	14:03
03 Room Air	0.000	14:04
04 Std. Gas	0.080	14:04
05 Room Air	0.000	14:05
06 Std. Gas	0.080	14:05
07 Room Air	0.000	14:06

Lot No = 264437  
Cyl No = 104  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

Dry cal check  
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      14:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:08
02 Std. Gas	0.079	14:08
03 Room Air	0.000	14:09
04 Std. Gas	0.080	14:09
05 Room Air	0.000	14:10
06 Std. Gas	0.079	14:10
07 Room Air	0.000	14:10

Lot No = 264437  
Cyl No = 104  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
#4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.14.00 09/16  
04/09/2026      14:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:16
02 Std. Gas	0.080	14:16
03 Room Air	0.000	14:17
04 Std. Gas	0.079	14:17
05 Room Air	0.000	14:18
06 Std. Gas	0.080	14:18
07 Room Air	0.000	14:19

Lot No = 264437  
Cyl No = 104  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
#7-9

Form 106-I8000