



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 2669 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 19825080A2 Cyl. # 15)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: WILL
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 930 psi Regulator: 950 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.025 AC
Sim. Ser #: MP3063
Lot #: 202509A
Exp. Date: 9/10/27
 Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 26-4437

Cylinder #: 104

Exp. Date: 1/29/28

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.080 Test 7 0.081

Test 2 0.080 Test 5 0.080 Test 8 0.081

Test 3 0.081 Test 6 0.080 Test 9 0.080

Average 0.080

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

Chris Lyngren
Inspector Signature

09 April 2026
Date

Janelle Pritchiller
Reviewer

10 Apr. 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 09:55

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19825080A2
Standard Cyl #? 15
Standard Expiration? 09/05/2027
Oper No? 133237

Flow Cal. Date: 06/14/2021
Slope 616
Intercept -441889

IR Calibration Date: 06/14/2021

	3um	9um
0th Coef(*100):	-27364	-15040
1st Coef(*100):	280250	139895
2nd Coef(*100):	444	862
H2O adj(mg/l*10k):	479	256

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 09:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:00
02 Std. Sol.	0.026	10:01
03 Room Air	0.000	10:01
04 Std. Sol.	0.026	10:02
05 Room Air	0.000	10:02
06 Std. Sol.	0.026	10:03
07 Room Air	0.000	10:04


08 Sim Temp = 34.0°C

Simul Ser No = MP3063

Std Sol No = 202509A

County = 08

Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *DOW AC check*
0.025AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 10:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:05
02 Std. Sol.	0.296	10:06
03 Room Air	0.000	10:06
04 Std. Sol.	0.298	10:07
05 Room Air	0.000	10:08
06 Std. Sol.	0.297	10:08
07 Room Air	0.000	10:09

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: High AC Check
0.300AC
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 10:12

Test	AC	Time
01 Room Air	0.000	10:13
02 Invalid Sample	X.XXX	10:13
03 Room Air	0.000	10:14

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Invalid sample -
Mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 10:14

Test	AC	Time
01 Diagnostic	OK	10:15
02 Room Air	RFI*	10:15
03 Room Air	0.000	10:16

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI Check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 10:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:16
02 Std. Gas	0.080	10:17
03 Room Air	0.000	10:17
04 Std. Gas	0.080	10:18
05 Room Air	0.000	10:18
06 Std. Gas	0.081	10:19
07 Room Air	0.000	10:19

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *Dry cal check
#1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 10:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:20
02 Std. Gas	0.080	10:20
03 Room Air	0.000	10:21
04 Std. Gas	0.080	10:21
05 Room Air	0.000	10:22
06 Std. Gas	0.080	10:22
07 Room Air	0.000	10:23

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry Cal Check
 #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/09/2026 10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:24
02 Std. Gas	0.081	10:24
03 Room Air	0.000	10:24
04 Std. Gas	0.081	10:25
05 Room Air	0.000	10:25
06 Std. Gas	0.080	10:26
07 Room Air	0.000	10:26

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#7-9

Form 106-I8000