



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Zack Moran 3414</i>		
Location <i>MCSD</i>	Intoxilyzer® Serial Number <i>80-003066</i>	
Gas Lot Number <i>26-4437</i>	Gas Cylinder Number <i>124</i>	Gas Expiration Date <i>01/29/2028</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.0*80* AC
 - B. 0.0*80* AC
 - C. 0.0*79* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>ZM #3414</i>	Date <i>04/11/2026</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i>	Date <i>28April2026</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = MCSO 8164.14.00 09/16
04/11/2026 03:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:03
02 Std. Gas	0.080	03:04
03 Room Air	0.000	03:04
04 Std. Gas	0.080	03:05
05 Room Air	0.000	03:05
06 Std. Gas	0.079	03:05
07 Room Air	0.000	03:06

Lot No = 264437
Cyl No = 124
Exp Date = 01/29/2028
County = 27 Oper No. = 131834



Operator Signature
ZACKARY MORAN

Remarks: *Test Pass*

Form 106-I8000