



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number 80-003066	Instrument Location MC50
-----------------------------------	------------------------------------

Reason for Install/Repair

Install After Receiving From Crime Laboratory Install After Location Change

Other (Specify) _____

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature [Signature] #3414	Date 04/11/2026
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen [Signature]	Certified Date 20April2026
---	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = MCSO 8164.14.00 09/16
04/11/2026 03:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 124
Standard Expiration? 01/29/2028
Oper No? 131834

Flow Cal. Date: 07/29/2010
Slope 806
Intercept -895033

IR Calibration Date: 07/29/2010
 3um 9um

0th Coef(*100): -14075 -5808
1st Coef(*100): 272113 138204
2nd Coef(*100): 1637 986
H2O adj(mg/l*10k): 289 148

***** Printer Test End *****



Operator Signature
ZACKARY MORAN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = MCSO 8164.14.00 09/16
04/11/2026 03:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:03
02 Std. Gas	0.080	03:04
03 Room Air	0.000	03:04
04 Std. Gas	0.080	03:05
05 Room Air	0.000	03:05
06 Std. Gas	0.079	03:05
07 Room Air	0.000	03:06

Lot No = 264437
Cyl No = 124
Exp Date = 01/29/2028
County = 27 Oper No. = 131834



Operator Signature
ZACKARY MORAN

Remarks: *Test Pass*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = MCSO 8164.14.00 09/16
04/11/2026 03:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:11
02 Std. Gas	0.080	03:11
03 Room Air	0.000	03:12
04 Std. Gas	0.080	03:12
05 Room Air	0.000	03:13
06 Std. Gas	RFI*	03:13
07 Room Air	0.000	03:13

*Invalid Test
Inhibited - RFI

Lot No = 264437
Cyl No = 124
Exp Date = 01/29/2028
County = 27 Oper No. = 131834



Operator Signature
ZACKARY MORAN

Remarks: RFI

Form 106-I8000