



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <i>80-007098</i>	Instrument Location <i>Troga PD (TI06)</i>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u><i>Annual Inspection (install)</i></u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> # <i>1133</i>	Date <i>4/20/26</i>
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Anna Ingemansen</i>	<i>Anna Ingemansen</i>	Certified Date <i>28 April 2026</i>
---	------------------------	--

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TIOG 8164.16.00 09/18
04/20/2026 18:57

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 250240803
Standard Cyl #? 54
Standard Expiration? 10/05/2026
Oper No? 131838

Flow Cal. Date: 09/26/2018
Slope 684
Intercept -588446

IR Calibration Date: 09/26/2018
 3um 9um

	3um	9um
0th Coef(*100):	-40948	-15551
1st Coef(*100):	237555	134140
2nd Coef(*100):	2571	1433
H2O adj(mg/l*10k):	1097	457

***** Printer Test End *****




Operator Signature
NICKOLAS OSTER

Remarks:

Print Test

I certify this is a true
and accurate copy.



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TIOG 8164.16.00 09/18
04/20/2026 19:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:03
02 Std. Gas	0.082	19:03
03 Room Air	0.000	19:04
04 Std. Gas	0.082	19:04
05 Room Air	0.000	19:04
06 Std. Gas	0.081	19:05
07 Room Air	0.000	19:05

Lot No = 250240803
Cyl No = 54
Exp Date = 10/05/2026
County = 53 Oper No. = 131838



Operator Signature
NICKOLAS OSTER

Remarks:

ACA Test

Form 106-I8000

I certify this is a true
and accurate copy.

X 

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TIOG 8164.16.00 09/18
04/20/2026 19:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	19:13
02 Room Air	0.000	19:13

*Invalid Test
Inhibited - RFI

Lot No = 250240803
Cyl No = 54
Exp Date = 10/05/2026
County = 53 Oper No. = 131838



Operator Signature
NICKOLAS OSTER

Remarks:

RFI Test

Form 106-I8000

I certify this is a true
and accurate copy.

X 