



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-00 5956	Instrument Location BMDC
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Install after Annual Inspection at Crime Lab	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): NA
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Dal M</i>	Date 4/20/26
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen <i>Anna Ingemansen</i>	Certified Date 30April2026
---	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = BMDC 8164.14.00 09/16
04/20/2026 16:28

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19825080A2
Standard Cyl #? 13
Standard Expiration? 09/05/2027
Oper No? 133598

Flow Cal. Date: 06/21/2013
Slope 706
Intercept -720901

IR Calibration Date: 04/04/2024

	3um	9um
0th Coef(*100):	-22945	-24559
1st Coef(*100):	278574	134068
2nd Coef(*100):	2387	1305
H2O adj(mg/l*10k):	467	490

***** Printer Test End *****


Operator Signature
DANIEL MAGELKY

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = BMDC 8164.14.00 09/16
04/20/2026 16:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:29
02 Std. Gas	0.077	16:30
03 Room Air	0.000	16:30
04 Std. Gas	0.077	16:31
05 Room Air	0.000	16:31
06 Std. Gas	0.076	16:32
07 Room Air	0.000	16:32

Lot No = 19825080A2
Cyl No = 13
Exp Date = 09/05/2027
County = 08 Oper No. = 133598



Operator Signature
DANIEL MAGELKY

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = BMDC 8164.14.00 09/16
04/20/2026 16:34

Test	AC	Time
01 Room Air	RFI*	16:36
02 Room Air	RFI*	16:36

*Invalid Test
Inhibited - RFI

Sub Name = TEST, RFI NA
Sub DOB = 01/01/2000
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 19825080A2
Cyl No = 13
Expiration Date = 09/05/2027
County = 08 Oper No. = 133598



Operator Signature
DANIEL MAGELKY

Remarks:

Form 106-I8000