



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Joshua S Gudvangen</i>		
Location <i>WARD</i>	Intoxilyzer® Serial Number <i>80-004191</i>	
Gas Lot Number <i>26-4437</i>	Gas Cylinder Number <i>126</i>	Gas Expiration Date <i>01/29/2028</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.071* AC
 - B. *0.671* AC
 - C. *0.671* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>4/21/2026</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i>	Date <i>28April2026</i>


Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

OMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = WARD 8154.14.00 09/16
04/21/2026 17:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:10
02 Std. Gas	0.079	17:11
03 Room Air	0.000	17:11
04 Std. Gas	0.079	17:12
05 Room Air	0.000	17:12
06 Std. Gas	0.079	17:13
07 Room Air	0.000	17:13

Lot No = 264437
Cyl. No = 126
Exp Date = 01/29/2028
County = 51 Open No. = 131334


Operator Signature
JOSHUA GUDANGEN

Remarks:

Form 106-18000