




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-004111	Instrument Location VARO
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  282	Date 04/21/2026
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen		Certified Date 28 April 2026
---	--	--

32 characters per line
abcdefghijklmnopqrstuvwxyz123456
ABCDEFGHIJKLMNOPQRSTUVWXYZ123456

42 characters per line
abcdefghijklmnopqrstuvwxyz1234567890abcdef
ABCDEFGHIJKLMNopqrstuvwxyz1234567890ABCDEF

INTOXILYZER 8000
Instrument Initialization
16:43 04/21/2026

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = WARD 8164.14.00 09/16
04/21/2026 17:09

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMNop


21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNopqrstu

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Reslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MenFull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.000
Standard Lot #? 264437
Standard Cyl #? 126
Standard Expiration? 01/29/2028
Oper No? 131334
Flow Cal. Date: 01/28/2010
Slope 738
Intercept -777574

IR Calibration Date: 06/25/2024
3um 9um

0th Coef(*100): -15752 -25149
1st Coef(*100): 278540 137807
2nd Coef(*100): 2188 1222
H2O adj(ng/1+10K): 202 388

***** Test End *****


Operator Signature
JOSHUA GUDUANGEN

Remarks:


Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = WARD 8164.14.00 09/16
04/21/2026 17:15

Test	AC	Time
01 Room Air	0.000	17:17
02 Reference	RFI*	17:17
03 Room Air	0.000	17:17

<Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST NA
Sub DOB = 01/01/2001
Sub Sex = Unknown Weight = NA
Test = DU1 Cyl = NA
Dr. Lic. = ND/NA
Lot No = 264437
Cyl No = 126
Expiration Date = 01/29/2028
County = 51 Oper No. = 131334


Operator Signature
JOSHUA GUDUANGEN

Remarks:


Form 106-18000

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = WARD 8164.14.00 09/16
04/21/2026 17:15

Test	AC	Time
01 Room Air	0.000	17:10
02 Std. Gas	0.079	17:11
03 Room Air	0.000	17:11
04 Std. Gas	0.079	17:12
05 Room Air	0.000	17:12
06 Std. Gas	0.079	17:13
07 Room Air	0.000	17:13

Lot No = 264437
Cyl No = 126
Exp Date = 01/29/2028
County = 51 Oper No. = 131334


Operator Signature
JOSHUA GUDUANGEN

Remarks:

Form 106-18000