



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4194 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 25024080A3 Cyl. # 2)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: KDPD
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 984 psi Regulator: 1000 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5290
Lot #: 202501E
Exp. Date: 1/28/27
 Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 32

Exp. Date: 9/5/27

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.078 Test 4 0.079 Test 7 0.079

Test 2 0.079 Test 5 0.079 Test 8 0.079

Test 3 0.079 Test 6 0.079 Test 9 0.079

Average 0.078

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

[Signature]
Inspector Signature

22 April 2026
Date

[Signature]
Reviewer

28 Apr. 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 25024080A3
Standard Cyl #? 2
Standard Expiration? 10/05/2026
Oper No? 133237

Flow Cal. Date: 06/23/2009
Slope 681
Intercept -622430

IR Calibration Date: 04/10/2024
 3um 9um

	3um	9um
0th Coef(*100):	-21912	-25573
1st Coef(*100):	266689	131809
2nd Coef(*100):	2490	1320
H2O adj(mg/l*10k):	376	518

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:21
02 Std. Sol.	0.020	08:23
03 Room Air	0.000	08:23
04 Std. Sol.	0.020	08:24
05 Room Air	0.000	08:24
06 Std. Sol.	0.020	08:25
07 Room Air	0.000	08:26

08 Sim Temp = 34.0°C

Simul Ser No = MP5290
Std Sol No = 202501E
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: LOW AC Check
 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:27
02 Std. Sol.	0.299	08:28
03 Room Air	0.000	08:28
04 Std. Sol.	0.299	08:29
05 Room Air	0.000	08:30
06 Std. Sol.	0.299	08:30
07 Room Air	0.000	08:31

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks:

High AC check
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:32

Test	AC	Time
01 Room Air	0.000	08:33
02 *Subject Test	INT*	08:33
03 Room Air	0.000	08:34

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 32
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent check
0.10AC + 0.05% Acetone

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:36

Test	AC	Time
01 Room Air	0.000	08:36
02 Invalid Sample	X.XXX	08:37
03 Room Air	0.000	08:37

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 32
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *invalid sample*
Mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:38

Test	AC	Time
01 Diagnostic	OK	08:39
02 Room Air	0.000	08:39
03 *Subject Test	RFI*	08:40
04 Room Air	0.000	08:40

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 32
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks:

RFI Check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:42
02 Std. Gas	0.078	08:42
03 Room Air	0.000	08:43
04 Std. Gas	0.079	08:43
05 Room Air	0.000	08:43
06 Std. Gas	0.079	08:44
07 Room Air	0.000	08:44

Lot No = 19825080A2
Cyl No = 32
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:46
02 Std. Gas	0.079	08:46
03 Room Air	0.000	08:47
04 Std. Gas	0.079	08:47
05 Room Air	0.000	08:47
06 Std. Gas	0.079	08:48
07 Room Air	0.000	08:48

Lot No = 19825080A2
Cyl No = 32
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
04/22/2026 08:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Gas	0.079	08:53
03 Room Air	0.000	08:53
04 Std. Gas	0.079	08:54
05 Room Air	0.000	08:54
06 Std. Gas	0.079	08:55
07 Room Air	0.000	08:55

Lot No = 19825080A2
Cyl No = 32
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#7-9

Form 106-I8000