



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-006498</b>	Instrument Location <b>Dunn Co Sheriff</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Yearly check at Crime Lab
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <b>Beth Bual</b>	Date <b>4-22-26</b>
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Anna Ingemansen</b>	<b>Anna Ingemansen</b>	Certified Date <b>29 April 2026</b>
---	------------------------	--

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006498  
Location = DUNN 8164.14.00 09/16  
04/22/2026 16:24

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 19825080A2  
Standard Cyl #? 14  
Standard Expiration? 09/05/2027  
Oper No? 131919

Flow Cal. Date: 08/18/2015  
Slope 669  
Intercept -637469

IR Calibration Date: 08/18/2015  
3um 9um

0th Coef(*100):	-28161	-22508
1st Coef(*100):	286657	143843
2nd Coef(*100):	2341	1281
H2O adj(mg/l*10k):	776	451

\*\*\*\*\* Printer Test End \*\*\*\*\*

Operator Signature  
BETH ISAAK

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006498  
Location = DUNN 8164.14.00 09/16  
04/22/2026 16:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Gas	0.080	16:26
03 Room Air	0.000	16:26
04 Std. Gas	0.078	16:27
05 Room Air	0.000	16:27
06 Std. Gas	0.079	16:28
07 Room Air	0.000	16:28

Lot No = 19825080A2

Cyl No = 14

Exp Date = 09/05/2027

County = 13

Oper No. = 131919



Operator Signature  
BETH ISAAK

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      - SN 80-006498  
Location = DUNN      8164.14.00 09/16  
04/22/2026      16:29

Test	AC	Time
01 Room Air	RFI*	16:33
02 Room Air	0.000	16:34

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST NA  
Sub DOB = 01/01/2001  
Sub Sex = Unknown      Weight = NA  
Test = OTH      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 19825080A2  
Cyl No = 14  
Expiration Date = 09/05/2027  
County = 13      Oper No. = 131919



Operator Signature  
BETH ISAAK

Remarks:

Form 106-I8000