




**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-005939</b>	Instrument Location <b>Williston LEC Visitation Room</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u><b>Annual Inspection</b></u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  <b>Cory Hubble #281</b>	Date <b>4/22/26</b>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	Certified Date
-----------------------	----------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005939  
Location = WILL      8164.14.00 09/16  
04/22/2026      10:29

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      19825080A2  
Standard Cyl #?      30  
Standard Expiration?      09/05/2027  
Oper No?      030922

Flow Cal. Date:      06/18/2020  
Slope      611  
Intercept      -592495

IR Calibration Date:      06/18/2020

	3um	9um
0th Coef(*100):	-15726	-19510
1st Coef(*100):	296627	143165
2nd Coef(*100):	-369	494
H2O adj(mg/l*10k):	220	240

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
-----  
Operator Signature  
COBY HUBBLE

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005939  
Location = WILL      8164.14.00 09/16  
04/22/2026      10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.079	10:30
03 Room Air	0.000	10:31
04 Std. Gas	0.079	10:31
05 Room Air	0.000	10:31
06 Std. Gas	0.080	10:32
07 Room Air	0.000	10:32

Lot No = 19825080A2  
Cyl No = 30  
Exp Date = 09/05/2027  
County = 53      Oper No. = 030922

  
\_\_\_\_\_  
Operator Signature  
COBY HUBBLE

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005939  
Location = WILL                              8164.14.00 09/16  
04/22/2026    10:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	10:34
02 Room Air	0.000	10:34

\*Invalid Test  
Inhibited - RFI

Lot No = 19825080A2  
Cyl No = 30  
Exp Date = 09/05/2027  
County = 53                                      Oper No. = 030922

 281  
\_\_\_\_\_  
Operator Signature  
GOBY HUBBLE

Remarks:

Form 106-I8000