



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7095 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or  No (If Yes, Lot # 08425080A1 Cyl. # 1)
  - Keys  Yes or  No
  - Power Cord  Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: OKPD
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or  MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 994 psi Regulator: 1000 psi  
Display and Regulator ± 50 psi of each other  Yes or  No  
Gas tank tare necessary?  Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5290  
Lot #: 202501E  
Exp. Date: 1/28/27  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 32

Exp. Date: 9/5/27

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS 9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.081 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

[Signature]  
Inspector Signature

23 April 2026  
Date

[Signature]  
Reviewer

28 Apr. 2026  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007095  
Location = TOXL                    8164.16.00 09/18  
04/23/2026                            10:15

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            08425080A1  
Standard Cyl #?            001  
Standard Expiration?        05/05/2027  
Oper No?                    133237

Flow Cal. Date:            09/27/2018  
Slope                        683  
Intercept                    -637452

IR Calibration Date:        06/21/2023

	3um	9um
0th Coef(*100):	-38728	-15234
1st Coef(*100):	250485	139595
2nd Coef(*100):	2553	1269
H2O adj(mg/l*10k):	921	383

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA INGEMANSEN

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
04/23/2026      10:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:16
02 Std. Sol.	0.020	10:17
03 Room Air	0.000	10:18
04 Std. Sol.	0.020	10:18
05 Room Air	0.000	10:19
06 Std. Sol.	0.020	10:19
07 Room Air	0.000	10:20

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 202501E  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: *Low AC Check*  
*0.020Ac*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
04/23/2026      10:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:21
02 Std. Sol.	0.296	10:22
03 Room Air	0.000	10:23
04 Std. Sol.	0.297	10:23
05 Room Air	0.000	10:24
06 Std. Sol.	0.297	10:25
07 Room Air	0.000	10:25

08 Sim Temp = 34.0°C

Simul Ser No = MP3003  
Std Sol No = 202408H  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: High AC check  
0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007095  
Location = TOXL                    8164.16.00 09/18  
04/23/2026                            10:27

Test	AC	Time
01 Room Air	0.000	10:29
02 *Subject Test	INT*	10:30
03 Room Air	0.000	10:30

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = OTH                          Cit = INTERFERENT  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 32  
Expiration Date = 09/05/2027  
County = 08                          Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: Interferent Check  
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007095  
Location = TOXL                8164.16.00 09/18  
04/23/2026                        10:33

Test	AC	Time
01 Room Air	0.000	10:34
02 Invalid Sample	X.XXX	10:34
03 Room Air	0.000	10:35

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                    Weight = NA  
Test = OTH                        Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 32  
Expiration Date = 09/05/2027  
County = 08                        Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

Invalid Sample  
Mouth Alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-007095  
Location = TOXL                    8164.16.00 09/18  
04/23/2026                            10:35

Test	AC	Time
01 Diagnostic	OK	10:36
02 Room Air	0.000	10:36
03 *Subject Test	RFI*	10:37
04 Room Air	0.000	10:37

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = OTH                                Cit = RFI  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 32  
Expiration Date = 09/05/2027  
County = 08                                Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA INGEMANSEN

Remarks: *RFI check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
04/23/2026      10:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:38
02 Std. Gas	0.080	10:38
03 Room Air	0.000	10:39
04 Std. Gas	0.080	10:39
05 Room Air	0.000	10:40
06 Std. Gas	0.080	10:40
07 Room Air	0.000	10:41

Lot No = 19825080A2  
Cyl No = 32  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007095  
Location = TOXL                    8164.16.00 09/18  
04/23/2026                            10:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:42
02 Std. Gas	0.081	10:42
03 Room Air	0.000	10:43
04 Std. Gas	0.080	10:43
05 Room Air	0.000	10:43
06 Std. Gas	0.080	10:44
07 Room Air	0.000	10:44

Lot No = 19825080A2  
Cyl No = 32  
Exp Date = 09/05/2027  
County = 08                            Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *Dry cal check  
#4-6*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
04/23/2026      10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Gas	0.080	10:45
03 Room Air	0.000	10:46
04 Std. Gas	0.080	10:46
05 Room Air	0.000	10:47
06 Std. Gas	0.080	10:47
07 Room Air	0.000	10:48

Lot No = 19825080A2  
Cyl No = 32  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

Dry cal check  
#7-9

Form 106-I8000