



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>80-007095</i>	Instrument Location <i>OKPD</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Michele B...</i>	Date <i>4/24/2026</i>
--------------------------------------------------	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen	<i>Anna Ingemansen</i>	Certified Date 29 April 2026
------------------------------------------	------------------------	---------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007095  
Location = OKPD                8164.16.00 09/18  
04/24/2026                      07:37

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                08425080A1  
Standard Cyl #?                1  
Standard Expiration?            05/05/2027  
Oper No?                        130690

Flow Cal. Date:                09/27/2018  
Slope                            683  
Intercept                       -637452

IR Calibration Date:            06/21/2023

	3um	9um
0th Coef(*100):	-38728	-15234
1st Coef(*100):	250485	139595
2nd Coef(*100):	2553	1269
H2O adj(mg/l*10k):	921	383

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
MATTHEW O'BRIEN

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = OKPD      8164.16.00 09/18  
04/24/2026      07:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:38
02 Std. Gas	0.079	07:39
03 Room Air	0.000	07:39
04 Std. Gas	0.079	07:40
05 Room Air	0.000	07:40
06 Std. Gas	0.080	07:40
07 Room Air	0.000	07:41

Lot No = 08425080A1

Cyl No = 1

Exp Date = 05/05/2027

County = 11

Oper No. = 130690



Operator Signature  
MATTHEW O'BRIEN

Remarks:


Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007095  
Location = OKPD                8164.16.00 09/18  
04/24/2026                        07:42

Test	AC	Time
01 Room Air	RFI*	07:44
02 Room Air	0.000	07:44

\*Invalid Test  
Inhibited - RFI



Operator Signature  
MATTHEW O'BRIEN

Remarks:

Form 106-I8000