



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6691 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 25024080 A 3 Cyl. # 9)
  - Keys  Yes or No
  - Power Cord  Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: NEWT
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 820 psi Regulator: 800 psi  
Display and Regulator ± 50 psi of each other  Yes or No  
Gas tank tare necessary?  Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5290  
Lot #: 202501E  
Exp. Date: 1128127  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 23

Exp. Date: 9/5/27

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.077 Test 4 0.078 Test 7 0.078

Test 2 0.077 Test 5 0.078 Test 8 0.078

Test 3 0.077 Test 6 0.078 Test 9 0.078

Average 0.077

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

Chris Lyman  
Inspector Signature

20 April 2026  
Date

Janella Putscheller  
Reviewer

22 Apr. 2026  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006691  
Location = TOXL                    8164.14.00 09/16  
04/20/2026                            12:08

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=-|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            25024080A3  
Standard Cyl #?            9  
Standard Expiration?       10/05/2026  
Oper No?                    133237

Flow Cal. Date:            08/28/2018  
Slope                        699  
Intercept                   -660610

IR Calibration Date:        04/03/2024  
                                  3um                    9um

	3um	9um
0th Coef(*100):	-12327	-23187
1st Coef(*100):	262018	135444
2nd Coef(*100):	2687	1323
H2O adj(mg/l*10k):	263	405

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:            *Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:13
02 Std. Sol.	0.021	12:14
03 Room Air	0.000	12:15
04 Std. Sol.	0.021	12:15
05 Room Air	0.000	12:16
06 Std. Sol.	0.020	12:17
07 Room Air	0.000	12:17


08 Sim Temp = 34.0°C

Simul Ser No = 202501E

Std Sol No = MP5290

County = 08

Oper No. = 133237



Operator Signature

ANNA INGEMANSEN

Remarks:

LOW AC Check

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:19
02 Std. Sol.	0.296	12:20
03 Room Air	0.000	12:20
04 Std. Sol.	0.297	12:21
05 Room Air	0.000	12:21
06 Std. Sol.	0.298	12:22
07 Room Air	0.000	12:23

08 Sim Temp = 34.0°C

Simul Ser No = 202408H

Std Sol No = MP3003

County = 08

Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

High AC Check  
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:25

Test	AC	Time
01 Room Air	0.000	12:26
02 *Subject Test	INT*	12:26
03 Room Air	0.000	12:27

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 23  
Expiration Date = 09/05/2027  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *Interferent Check*  
*0.10AC + 0.05% Acetone*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:28

Test	AC	Time
01 Room Air	0.000	12:28
02 Invalid Sample	X.XXX	12:29
03 Room Air	0.000	12:30

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 23  
Expiration Date = 09/05/2027  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

Invalid sample  
Mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:31

Test	AC	Time
01 Diagnostic	OK	12:32
02 Room Air	RFI*	12:32
03 Room Air	0.000	12:33

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 23  
Expiration Date = 09/05/2027  
County = 08      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:37
02 Std. Gas	0.077	12:37
03 Room Air	0.000	12:38
04 Std. Gas	0.077	12:38
05 Room Air	0.000	12:39
06 Std. Gas	0.077	12:39
07 Room Air	0.000	12:39

Lot No = 19825080A2

Cyl No = 23

Exp Date = 09/05/2027

County = 08

Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: Dry Cal check  
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:55
02 Std. Gas	0.078	12:56
03 Room Air	0.000	12:56
04 Std. Gas	0.078	12:56
05 Room Air	0.000	12:57
06 Std. Gas	0.078	12:57
07 Room Air	0.000	12:58

Lot No = 19825080A2  
Cyl No = 23  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: *Dry cal check  
#4-6*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = TOXL      8164.14.00 09/16  
04/20/2026      12:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Gas	0.078	12:59
03 Room Air	0.000	13:00
04 Std. Gas	0.078	13:00
05 Room Air	0.000	13:00
06 Std. Gas	0.078	13:01
07 Room Air	0.000	13:01

Lot No = 19825080A2  
Cyl No = 23  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
#7-a

Form 106-I8000