



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <u>25024080A3 80-006681 aei 5/8/26</u>	Instrument Location <u>CARS</u>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): annual inspection
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

aei
5/8/26

Field Inspector Signature 	Date <u>4-29-26</u>
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <u>Anna Ingemansen</u>	Certified Date <u>08 May 2026</u>
---	--------------------------------------

Serial number listed on the print test, ACA test and RFI test is 80-006681. The Form 104-G, Print test and ACA test and RFI test were sent to the crime Laboratory. This does not affect the installation of the instrument. aei 5/8/26

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = CARS 8164.14.00 09/16
04/29/2026 08:46

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 25024080A3
Standard Cyl #? 49
Standard Expiration? 10/05/2026
Oper No? 132318

Flow Cal. Date: 05/27/2020
Slope 651
Intercept -622365

IR Calibration Date: 05/27/2020
 3um 9um

0th Coef(*100): -10420 -19573
1st Coef(*100): 269401 137403
2nd Coef(*100): 1095 923
H2O adj(mg/l*10k): 127 287

***** Printer Test End *****



Operator Signature
TERRY DILLON

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = CARS 8164.14.00 09/16
04/29/2026 08:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:58
02 Std. Gas	0.079	08:58
03 Room Air	0.000	08:59
04 Std. Gas	0.079	08:59
05 Room Air	0.000	09:00
06 Std. Gas	0.080	09:00
07 Room Air	0.000	09:01

Lot No = 25024080A3
Cyl No = 49
Exp Date = 10/05/2026
County = 19 Oper No. = 132318


Operator Signature
TERRY DILLON

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = CARS 8164.14.00 09/16
04/29/2026 09:02

Test	AC	Time
01 Room Air	0.000	09:04
02 *Subject Test	RFI*	09:05
03 Room Air	0.000	09:05

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/01/2000
Sub Sex = Male Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 25024080A3
Cyl No = 49
Expiration Date = 10/05/2026
County = 19 Oper No. = 132318


Operator Signature
TERRY DILLON

Remarks:

Form 106-I8000