



**NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION**

**INTOXILYZER® 8000 ANNUAL INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-00 6508 Inspection Location: TOXL

**A. Pre-Inspection**

1. Items with Instrument:
  - Gas Cylinder Yes or No (If Yes, Lot # \_\_\_\_\_ Cyl. # \_\_\_\_\_)
  - Keys Yes or No
  - Power Cord Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: TOXL
5.  Battery Check
  - Was the external battery pack replaced? Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or No
  - Replaced Breath Tube O-Ring Yes or No

**B. General Setup and Checks:**

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 853 psi Regulator: 850 psi  
Display and Regulator ± 50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

**C. Tests (Sign and attach test records):**

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5290  
Lot #: 202501E  
Exp. Date: 1/28/27  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 20

Exp. Date: 9/5/27

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS 9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.079      Test 4 0.079      Test 7 0.079

Test 2 0.079      Test 5 0.078      Test 8 0.079

Test 3 0.079      Test 6 0.079      Test 9 0.079

Average 0.078

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

Alex Lyman  
Inspector Signature

07 May 2026  
Date

Janelle Putschman  
Reviewer

07 May 2026  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006508  
Location = TOXL                 8164.14.00 09/16  
05/07/2026                        08:54

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=-|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 02621080A1  
Standard Cyl #?                 6  
Standard Expiration?           03/05/2023  
Oper No?                        133237

Flow Cal. Date:                 08/20/2015  
Slope                            644  
Intercept                        -568181

IR Calibration Date:            08/20/2015  
                                    3um                            9um

-----  
0th Coef(\*100):                -22365                        -18627  
1st Coef(\*100):                256969                        136095  
2nd Coef(\*100):                3168                            1444  
H2O adj(mg/l\*10k):            693                            453

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA INGEMANSEN

Remarks:                        Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006508  
Location = TOXL      8164.14.00 09/16  
05/07/2026      08:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:55
02 Std. Sol.	0.021	08:57
03 Room Air	0.000	08:58
04 Std. Sol.	0.021	08:59
05 Room Air	0.000	08:59
06 Std. Sol.	0.021	09:00
07 Room Air	0.000	09:01

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 202501E  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:      *Low AC Check*  
                         *0.020 AC check aei*  
                         *5/7/26*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006508  
Location = TOXL      8164.14.00 09/16  
05/07/2026      09:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:02
02 Std. Sol.	0.293	09:03
03 Room Air	0.000	09:04
04 Std. Sol.	0.295	09:05
05 Room Air	0.000	09:05
06 Std. Sol.	0.295	09:06
07 Room Air	0.000	09:06

08 Sim Temp = 34.0°C

Simul Ser No = MP3003  
Std Sol No = 202408H  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: High AC check  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006508  
Location = TOXL                                      8164.14.00 09/16  
05/07/2026    09:14

Test	AC	Time
01 Room Air	0.000	09:15
02 *Subject Test	INT*	09:17
03 Room Air	0.000	09:18

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                      Weight = NA  
Test = OTH                                      Cit = INTERFERENT  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 20  
Expiration Date = 09/05/2027  
County = 08                                      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

interferent check  
0.10 AC + 0.05% <sup>aei</sup> 5/7/26  
Acetone

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006508  
Location = TOXL                    8164.14.00 09/16  
05/07/2026                            09:34

Test	AC	Time
01 Room Air	0.000	09:35
02 Invalid Sample	X.XXX	09:35
03 Room Air	0.000	09:36

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = OTH                            Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 20  
Expiration Date = 09/05/2027  
County = 08                            Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: Invalid Sample  
Mouth Alcohol

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006508  
Location = TOXL      8164.14.00 09/16  
05/07/2026      09:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:40
02 Std. Gas	0.079	09:40
03 Room Air	0.000	09:41
04 Std. Gas	0.079	09:41
05 Room Air	0.000	09:42
06 Std. Gas	0.079	09:42
07 Room Air	0.000	09:43

Lot No = 19825080A2  
Cyl No = 20  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry Cal check  
# 1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006508  
Location = TOXL      8164.14.00 09/16  
05/07/2026      09:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:47
02 Std. Gas	0.079	09:47
03 Room Air	0.000	09:48
04 Std. Gas	0.078	09:48
05 Room Air	0.000	09:48
06 Std. Gas	0.079	09:49
07 Room Air	0.000	09:49

Lot No = 19825080A2  
Cyl No = 20  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006508  
Location = TOXL      8164.14.00 09/16  
05/07/2026      09:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:50
02 Std. Gas	0.079	09:51
03 Room Air	0.000	09:51
04 Std. Gas	0.079	09:52
05 Room Air	0.000	09:52
06 Std. Gas	0.079	09:52
07 Room Air	0.000	09:53

Lot No = 19825080A2  
Cyl No = 20  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

Form 106-I8000