



**NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION**

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6678 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 26-4437 Cyl. # 123)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: GFPD
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 857 psi Regulator: 900 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5290
Lot #: 202501E
Exp. Date: 1/28/27
 Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 25024080A3

Cylinder #: 1

Exp. Date: 10/15/26

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS 9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.081 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.080 Test 8 0.081

Test 3 0.082 Test 6 0.080 Test 9 0.081

Average 0.080

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

[Signature]
Inspector Signature

12 May 2026
Date

[Signature]
Reviewer

14 May 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:10

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 123
Standard Expiration? 01/29/2028
Oper No? 133237

Flow Cal. Date: 08/22/2016
Slope 686
Intercept -611596

IR Calibration Date: 08/22/2016

	3um	9um
0th Coef(*100):	-15445	-21650
1st Coef(*100):	251854	131766
2nd Coef(*100):	3094	1378
H2O adj(mg/l*10k):	549	513

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:12
02 Std. Sol.	0.021	10:13
03 Room Air	0.000	10:14
04 Std. Sol.	0.021	10:14
05 Room Air	0.000	10:15
06 Std. Sol.	0.021	10:16
07 Room Air	0.000	10:16

08 Sim Temp = 34.0°C

Simul Ser No = MP5290
Std Sol No = 202501E
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *Low AC check*
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Sol.	0.292	10:19
03 Room Air	0.000	10:19
04 Std. Sol.	0.291	10:20
05 Room Air	0.000	10:21
06 Std. Sol.	0.291	10:21
07 Room Air	0.000	10:22

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *High Ac check*
 0.300Ac

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:29

Test	AC	Time
01 Room Air	0.000	10:30
02 *Subject Test	INT*	10:30
03 Room Air	0.000	10:31

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 25024080A3
Cyl No = 1
Expiration Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *interferent check*
0.10AC + 0.05% Acetone

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:37
02 Std. Gas	0.081	10:38
03 Room Air	0.000	10:38
04 Std. Gas	0.080	10:39
05 Room Air	0.000	10:39
06 Std. Gas	0.082	10:39
07 Room Air	0.000	10:40

Lot No = 25024080A3
Cyl No = 1
Exp Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *dry cal check*
 #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:41
02 Std. Gas	0.081	10:42
03 Room Air	0.000	10:42
04 Std. Gas	0.080	10:42
05 Room Air	0.000	10:43
06 Std. Gas	0.080	10:43
07 Room Air	0.000	10:44

Lot No = 25024080A3
Cyl No = 1
Exp Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry calcheck
 #46

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = TOXL 8164.14.00 09/16
05/12/2026 10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Gas	0.081	10:45
03 Room Air	0.000	10:46
04 Std. Gas	0.081	10:46
05 Room Air	0.000	10:47
06 Std. Gas	0.081	10:47
07 Room Air	0.000	10:48

Lot No = 25024080A3
Cyl No = 1
Exp Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry Cal check
#7-9

Form 106-I8000