



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 0512 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or  No (If Yes, Lot # 26-4437 Cyl. # 120)
  - Keys  Yes or  No
  - Power Cord  Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: GFSO
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring  Yes or  No
  - Replaced Breath Tube O-Ring  Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or  MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 886 psi Regulator: 900 psi  
Display and Regulator ± 50 psi of each other  Yes or  No  
Gas tank tare necessary?  Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5290  
Lot #: 202501E  
Exp. Date: 1/28/27  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 20

Exp. Date: 9/5/27

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1CS9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.078 Test 4 0.079 Test 7 0.080

Test 2 0.078 Test 5 0.079 Test 8 0.079

Test 3 0.079 Test 6 0.080 Test 9 0.080

Average 0.079

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

[Signature]  
Inspector Signature

14 May 2026  
Date

[Signature]  
Reviewer

15 May 2026  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006512  
Location = TOXL                    8164.14.00 09/16  
05/14/2026                            08:21

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            264437  
Standard Cyl #?            120  
Standard Expiration?       01/29/2028  
Oper No?                    133237

Flow Cal. Date:            08/20/2015  
Slope                      680  
Intercept                  -660856

IR Calibration Date:       08/21/2015  
                                  3um                    9um

-----  
0th Coef(\*100):            -23051                -19261  
1st Coef(\*100):            266836                137135  
2nd Coef(\*100):            3974                   1466  
H2O adj(mg/l\*10k):        785                    487

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA INGEMANSEN

Remarks:            *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006512  
Location = TOXL      8164.14.00 09/16  
05/14/2026      08:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:27
02 Std. Sol.	0.021	08:28
03 Room Air	0.000	08:28
04 Std. Sol.	0.021	08:29
05 Room Air	0.000	08:29
06 Std. Sol.	0.021	08:30
07 Room Air	0.000	08:31

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 202501E  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:      *Low AC check*  
                    *0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006512  
Location = TOXL      8164.14.00 09/16  
05/14/2026      08:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:32
02 Std. Sol.	0.300	08:33
03 Room Air	0.000	08:34
04 Std. Sol.	0.300	08:35
05 Room Air	0.000	08:35
06 Std. Sol.	0.300	08:36
07 Room Air	0.000	08:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3003  
Std Sol No = 202408H  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: High Ac Check  
0.300Ac

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006512  
Location = TOXL                    8164.14.00 09/16  
05/14/2026                            08:43

Test	AC	Time
01 Room Air	0.000	08:43
02 Invalid Sample	X.XXX	08:43
03 Room Air	0.000	08:44

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = OTH                          Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 20  
Expiration Date = 09/05/2027  
County = 08                          Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:    *invalid sample*  
                  *mouth alcohol*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006512  
Location = TOXL                              8164.14.00 09/16 \*  
05/14/2026    08:45

Test	AC	Time
01 Diagnostic	OK	08:46
02 Room Air	0.000	08:46
03 *Subject Test	RFI*	08:46
04 Room Air	0.000	08:47

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                              Weight = NA  
Test = OTH                                      Cit = RFI  
Dr. Lic. = ND/TES989643  
Lot No = 19825080A2  
Cyl No = 20  
Expiration Date = 09/05/2027  
County = 08                              Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA INGEMANSEN

Remarks:                      RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006512  
Location = TOXL                    8164.14.00 09/16  
05/14/2026                            08:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:48
02 Std. Gas	0.078	08:48
03 Room Air	0.000	08:49
04 Std. Gas	0.078	08:49
05 Room Air	0.000	08:50
06 Std. Gas	0.079	08:50
07 Room Air	0.000	08:50

Lot No = 19825080A2  
Cyl No = 20  
Exp Date = 09/05/2027  
County = 08                            Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:    *Dry cal check*  
                  *#1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006512  
Location = TOXL      8164.14.00 09/16  
05/14/2026      08:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:54
02 Std. Gas	0.079	08:55
03 Room Air	0.000	08:55
04 Std. Gas	0.079	08:56
05 Room Air	0.000	08:56
06 Std. Gas	0.080	08:56
07 Room Air	0.000	08:57

Lot No = 19825080A2  
Cyl No = 20  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:      Dry cal check  
                         #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006512  
Location = TOXL      8164.14.00 09/16  
05/14/2026      08:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:58
02 Std. Gas	0.080	08:59
03 Room Air	0.000	08:59
04 Std. Gas	0.079	08:59
05 Room Air	0.000	09:00
06 Std. Gas	0.080	09:00
07 Room Air	0.000	09:01

Lot No = 19825080A2  
Cyl No = 20  
Exp Date = 09/05/2027  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
#7-9

Form 106-I8000