



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5958 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder Yes or No (If Yes, Lot # \_\_\_\_\_ Cyl. # \_\_\_\_\_)
  - Keys Yes or No
  - Power Cord Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: TOXL
5.  Battery Check
  - Was the external battery pack replaced? Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or No
  - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 800 psi Regulator: 841 psi  
Display and Regulator ± 50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5290  
Lot #: 202501E  
Exp. Date: 4/28/26 del 5/27/26 11/28/27  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 25024080A3

Cylinder #: 59

Exp. Date: 10/5/26

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.078      Test 4 0.079      Test 7 0.079

Test 2 0.078      Test 5 0.078      Test 8 0.079

Test 3 0.078      Test 6 0.080      Test 9 0.080

Average 0.078

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

Oliver Lyman  
Inspector Signature

27 May 2026  
Date

Janelle Putschiller  
Reviewer

28 May 2026  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005958  
Location = TOXL                    8164.14.00 09/16  
05/27/2026                            13:33

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34921080A1  
Standard Cyl #?            60  
Standard Expiration?       02/05/2024  
Oper No?                    133237

Flow Cal. Date:            06/21/2013  
Slope                        716  
Intercept                   -760436

IR Calibration Date:        04/07/2015  
                                  3um                            9um

-----  
0th Coef(\*100):            -17179                        -17694  
1st Coef(\*100):            259328                        130890  
2nd Coef(\*100):            3473                            1654  
H2O adj(mg/l\*10k):        425                            422

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA INGEMANSEN

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = TOXL      8164.14.00 09/16  
05/27/2026      13:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Sol.	0.019	13:36
03 Room Air	0.000	13:37
04 Std. Sol.	0.020	13:37
05 Room Air	0.000	13:38
06 Std. Sol.	0.020	13:39
07 Room Air	0.000	13:39

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 202501E  
County = 08      Oper No. = 133237

Operator Signature  
ANNA INGEMANSEN

Remarks: Low Ae check  
0.020Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = TOXL      8164.14.00 09/16  
05/27/2026      13:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:41
02 Std. Sol.	0.297	13:43
03 Room Air	0.000	13:44
04 Std. Sol.	0.297	13:44
05 Room Air	0.000	13:45
06 Std. Sol.	0.297	13:46
07 Room Air	0.000	13:46


08 Sim Temp = 34.0°C

Simul Ser No = MP3003

Std Sol No = 202408H

County = 08

Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

High AC check  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = TOXL                8164.14.00 09/16  
05/27/2026                        14:04

Test	AC	Time
01 Room Air	0.000	14:05
02 *Subject Test	INT*	14:05
03 Room Air	0.000	14:06

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                    Weight = NA  
Test = OTH                        Cit = INTERFERENT  
Dr. Lic. = ND/TES989643  
Lot No = 25024080A3  
Cyl No = 59  
Expiration Date = 10/05/2026  
County = 08                        Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: *interferent check*  
*0.10AC + 0.05% Acetone*  
Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = TOXL                8164.14.00 09/16  
05/27/2026                        14:13

Test	AC	Time
01 Room Air	0.000	14:14
02 Invalid Sample	X.XXX	14:15
03 Room Air	0.000	14:15

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                    Weight = NA  
Test = OTH                        Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 25024080A3  
Cyl No = 59  
Expiration Date = 10/05/2026  
County = 08                        Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: *invalid sample  
mouth alcohol*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = TOXL                8164.14.00 09/16  
05/27/2026                      14:16

Test	AC	Time
01 Diagnostic	OK	14:16
02 Room Air	0.000	14:17
03 *Subject Test	RFI*	14:18
04 Room Air	0.000	14:18

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = OTH                              Cit = RFI  
Dr. Lic. = ND/TES989643  
Lot No = 25024080A3  
Cyl No = 59  
Expiration Date = 10/05/2026  
County = 08                              Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:    *RFI check*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = TOXL                    8164.14.00 09/16  
05/27/2026                            14:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:19
02 Std. Gas	0.078	14:20
03 Room Air	0.000	14:20
04 Std. Gas	0.078	14:21
05 Room Air	0.000	14:21
06 Std. Gas	0.078	14:21
07 Room Air	0.000	14:22

Lot No = 25024080A3  
Cyl No = 59  
Exp Date = 10/05/2026  
County = 08                            Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: *Dry cal check  
#1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = TOXL      8164.14.00 09/16  
05/27/2026      14:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:23
02 Std. Gas	0.079	14:23
03 Room Air	0.000	14:24
04 Std. Gas	0.078	14:24
05 Room Air	0.000	14:25
06 Std. Gas	0.080	14:25
07 Room Air	0.000	14:26

Lot No = 25024080A3  
Cyl No = 59  
Exp Date = 10/05/2026  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA INGEMANSEN

Remarks: Dry cal check  
#4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = TOXL      8164.14.00 09/16  
05/27/2026      14:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:39
02 Std. Gas	0.079	14:39
03 Room Air	0.000	14:40
04 Std. Gas	0.079	14:40
05 Room Air	0.000	14:40
06 Std. Gas	0.080	14:41
07 Room Air	0.000	14:41

Lot No = 25024080A3  
Cyl No = 59  
Exp Date = 10/05/2026  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

Dry cal check  
#7-9

Form 106-I8000