



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6688 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: TOXL UNDP aei 22May26
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
 - Display: 826 psi Regulator: 850 psi
 - Display and Regulator ± 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value ≤ 0.03 AC: 0.020 AC
 - Sim. Ser #: MP5290
 - Lot #: 202501E
 - Exp. Date: 1/28/27
 - Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 25024080A3

Cylinder #: 1

Exp. Date: 10/5/26

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.078 Test 4 0.078 Test 7 0.078

Test 2 0.078 Test 5 0.078 Test 8 0.079

Test 3 0.078 Test 6 0.079 Test 9 0.079

Average 0.078

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

Ann Lyman
Inspector Signature

22 May 2026
Date

Janelle Putschler
Reviewer

28 May 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 10:28

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19825080A2
Standard Cyl #? 31
Standard Expiration? 09/05/2027
Oper No? 133237

Flow Cal. Date: 09/07/2016
Slope 698
Intercept -642832

IR Calibration Date: 04/04/2024
 3um 9um

0th Coef(*100): -28173 -24020
1st Coef(*100): 268738 133278
2nd Coef(*100): 2318 1374
H2O adj(mg/l*10k): 506 480

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 10:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Sol.	0.290	10:44
03 Room Air	0.000	10:44
04 Std. Sol.	0.291	10:45
05 Room Air	0.000	10:46
06 Std. Sol.	0.291	10:46
07 Room Air	0.000	10:47

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: High AC Check
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 10:48

Test	AC	Time
01 Room Air	0.000	10:49
02 *Subject Test	INT*	10:51
03 Room Air	0.000	10:52

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 25024080A3
Cyl No = 1
Expiration Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent Check
0.10AC + 0.05% Acetone
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 10:54

Test	AC	Time
01 Room Air	0.000	10:55
02 Invalid Sample	X.XXX	10:55
03 Room Air	0.000	10:56

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALIDIS
Dr. Lic. = ND/TES989643
Lot No = 25024080A3
Cyl No = 1
Expiration Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks:

Invalid Sample
Mouth Alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 10:56

Test	AC	Time
01 Diagnostic	OK	10:57
02 Room Air	RFI*	10:57
03 Room Air	0.000	10:57

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI
Dr. Lic. = ND/TES989643
Lot No = 25024080A3
Cyl No = 1
Expiration Date = 10/05/2026
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 10:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:58
02 Std. Gas	0.078	10:59
03 Room Air	0.000	10:59
04 Std. Gas	0.078	10:59
05 Room Air	0.000	11:00
06 Std. Gas	0.078	11:00
07 Room Air	0.000	11:01

Lot No = 25024080A3
Cyl No = 1
Exp Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: *Dry cal check*
 #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 11:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:02
02 Std. Gas	0.078	11:02
03 Room Air	0.000	11:03
04 Std. Gas	0.078	11:03
05 Room Air	0.000	11:04
06 Std. Gas	0.079	11:04
07 Room Air	0.000	11:04

Lot No = 25024080A3
Cyl No = 1
Exp Date = 10/05/2026
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/22/2026 11:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:06
02 Std. Gas	0.078	11:06
03 Room Air	0.000	11:06
04 Std. Gas	0.079	11:07
05 Room Air	0.000	11:07
06 Std. Gas	0.079	11:08
07 Room Air	0.000	11:08

Lot No = 25024080A3
Cyl No = 1
Exp Date = 10/05/2026
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#7-9

Form 106-I8000